

Member State of OIML
United Kingdom of Great Britain
and Northern Ireland

OIML Certificate No R117/1995-GB1-13.01

OIML CERTIFICATE OF CONFORMITY

Issuing authority: National Measurement Office

Person responsible: Paul Dixon – Product Certification Manager

Applicant: Pumptronics Europe Ltd

Address: Folgate Road
North Walsham

Norwich

Manufacturer: Same as applicant

Address:

Identification of the

certified pattern: Liquids other than water dispenser, C Series

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organisation of Legal Metrology (OIML):

OIML R117 - Edition 1995(E) for accuracy class: 0.5

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the certificates reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.

Issue Date: 18 December 2013

Reference No: T1117/0045

Signatory: G Stones

for Chief Executive





The conformity was established by tests described in the associated pattern evaluation report P00529.

Characteristics:

Table 1 -

Petrol, ethanol blended petrol, petrol with up to 5% methanol, petrol with up to 15% MTBE, diesel and biodiesel blends.

Accuracy Class	0.5
Mechanical Environment Class	M1
Electromagnetic Environment	E1
Class	
V_{min}	= 2 litres if Q _{max} ≤ 90 lpm
	= 5 litres if Q _{max} > 90 lpm
	often determined by national re-
	verification practices.

Note: Where two (90 lpm) pumps and two meters are combined into a single outlet flowrates up to 180 LPM are approved.

 Table 2
 Prime Component Summary

Manufacturer	Туре	Evaluation Certificate		
Meters				
Tatsuno	FM1007	GB-1502		
Pumptronics	FM-00H101	GB-1503		
Gas Separators				
Tatsuno	FP1012	GB-1506		
Tatsuno	FP1001	GB-1505		
Pumptronics	FP-00H102	GB-1504		
Electronic Calculator				
Pumptronics	MCU	GB-1507		

Prime Component Characteristics Table 3

	Ambient	Liquid	Viscosity	Q _{min}	Q _{max}	P _{max}
	temp	temp	range			
FM1007	-25°C to	-10°C to	0.4mPa.s to	3 lpm	90 lpm	3.5 bar
meter	+55°C	+40°C	17 mPa.s			
FM-00H101	-25°C to	-10°C to	0.4mPa.s to	3 lpm	90 lpm	3.5 bar
meter	+55°C	+40°C	17 mPa.s			
PTF Adblue	-25°C to	-10°C to	-	4 lpm	40 lpm	3.5 bar
	+55°C	+50°C				
FP1012	-25°C to	-10°C to	0.4 mPa.s to	-	90 lpm*	3.5 bar
	+55°C	+40°C	17 mPa.s			
FP1001	-25°C to	-10°C to	0.4 mPa.s to	-	90 lpm**	3.5 bar
	+55°C	+40°C	17 mPa.s			
FP-00H102	-25°C to	-10°C to	0.4 mPa.s to	-	90 lpm**	3.5 bar
	+55°C	+40°C	17 mPa.s			
MCU	-25°C to	-	-	_	-	-
(calculator)	+55°C					

^{*} Gas indicator (sight glass) required at flow-rates above 40 lpm ** Gas indicator (sight glass) required at all flow-rates

Certificate History

ISSUE NO.	DATE	DESCRIPTION
R117 1995-GB1-13.01	18 December 2013	Certificate first issued