


OIML Member State
Japan



OIML Certificate N°
R117/1995-JP1-02.01
Revision 1

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: National Metrology Institute of Japan /National Institute of
Advanced Industrial Science and Technology (NMIJ/AIST)
Address: AIST Tsukuba Central 3-9
Tsukuba Ibaraki 305-8563, Japan
Person responsible: Dr. Hiroyuki Yoshikawa, President of AIST

Applicant

Name: TATSUNO CORPORATION
Address: 2-12-13, Shibaura Minato-ku, TOKYO, 108-8520, Japan

Manufacturer of the certified type

Name: TATSUNO CORPORATION Yokohama plant
Address: 200, Iijima-cho, Sakae-ku, Yokohama 244-8501, Japan

Identification of the certified type:

Fuel dispenser for motor vehicles, Tatsuno NEO series
Further characteristics see page 2 to 4

This certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R117
Edition 1995
for accuracy class 0.5

R118
Edition 1995


This certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This certificate does not bestow any form of legal international approval.

OIML Member State
Japan



OIML Certificate N°
R117/1995-JP1-02.01
Revision 1



The conformity was established by tests described in the associated test report N° R117/1995-JP-2006-01, that includes 30 pages.

The Issuing Authority

The OIML Member

NMIJ/AIST



Dr. H. Yoshida

Dr. Y. Miki

President of AIST

2006-09-13

2006-09-13

Characteristics: Fuel dispenser for motor vehicles, Tatsuno NEO series

| Accuracy class | Maximum flowrate (Qmax) | Minimum flowrate (Qmin) | Minimum measured Quantity (litres) | Maximum volume Indication (number of digits) | Maximum unit price (number of digits) | Maximum price-to-pay (number of digits) |
|----------------|-------------------------|-------------------------|------------------------------------|--|---------------------------------------|---|
| Low flowrate | | | | | | |
| 0.5 | 45 L/min | 3 L/min | 2 | 6 | 4 | 6 |
| High flowrate | | | | | | |
| 0.5 | 90 L/min | 3 L/min | 2 | 6 | 4 | 6 |

Liquids: Petrol, Kerosene, Diesel

Comprising of:

- Meter

Manufacturer: TATSUNO CORPORATION
Pattern designation: FM-1007

- Gas elimination device (integral with pump)

Manufacturer: TATSUNO CORPORATION
Pattern designation: FP-1001



OIML Member State
Japan



OIML Certificate N°
R117/1995-JP1-02.01
Revision 1

- Measuring transducer

Manufacturer: TATSUNO CORPORATION
Pattern designation: EK-1025

- Calculator


Manufacturer: TATSUNO CORPORATION
Pattern designation: ER-1161, ER-1162, ER-1163, ER-1164, ER-1165, ER-1166

- Hose

Manufacturer: TATSUNO CORPORATION
Pattern designation: ZH-1140, ZH-1195, ZH-1142, ZH-1203, ZH-1141, ZH-1196,
ZH-1143, ZH-1204, ZH-1107

Manufacturer: ELAFLEX
Pattern designation: SLIMLINE16, SLIMLINE19, SLIMLINE25, COAX

Manufacturer: Goodyear
Pattern designation: FLEXSTEEL HARDWALL 5/8, FLEXSTEEL HARDWALL
3/4, FLEXSTEEL HARDWALL 1, 2BRADE BC SOFTWALL 5/8, 2BRADE BC
SOFTWALL 3/4, FLEXSTEEL VAPOR ASSIST



OIML Member State
Japan



OIML Certificate N°
R117/1995-JP1-02.01
Revision 1

• Nozzle

Manufacturer: TATSUNO CORPORATION
Pattern designation: FN-1001, FN-1002, FN-1021, FN-1017

Manufacturer: ELAFLEX
Pattern designation: ZVA Slimline, ZVA 25, ZVA 200-GR

Manufacturer: OPW
Pattern designation: TYPE 11, TYPE 7, TYPE 11VAI

Manufacturer: EMCO
Pattern designation: A2500, A2501, A4500, A4505

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.