



**FEDERAL AGENCY FOR TECHNICAL
REGULATION AND METROLOGY OF RUSSIA**

OIML Member State
The Russian Federation

OIML Certificate №
R117/1995-RU1-06.07

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: VNIIMS
Address: 46, Ozernaya str, G-361, Moscow, 119361,
the Russian Federation

Person responsible: Dr. S.A. Kononogov

Applicant

Name: CLOSED JOINT-STOCK COMPANY «NARA»
Address: 1, Polevaya, Serpukhov, Moscow region, 142 207,
Russia

Manufacturer of the certified pattern is the Applicant

Identification of the certified pattern: Nara 28V

Further characteristics are set out on page 2

This certificate attests the conformity of the above-mentioned pattern (represented by the sample identified in the associated test reports) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R117
edition 1995
for accuracy class 0.5

R118
edition 1995

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.
This certificate does not bestow any form of legal international approval.
The conformity was established by tests described in associated test report number which includes 45 pages.

The issuing authority

The CIML member



Dr. S.A. Kononogov
Director
Date: 10.07.2006

Dr. L. K. Issaev
Professor
Date: 10.07.2006

Characteristics: Fuel Dispensers Nara 28V comprising of one or more system in every enclosure and each sub system consist of the following:

1. Piston flow meter type
2. Gear and vane pump with gas eliminator.
3. Calculator-Electronic Register Assembly
4. Encoder Measuring transducer

Accuracy class	Maximum flowrate l/min	Minimum flowrate l/min	Minimum measured quantity 1	Temperature range °C	Maximum unit price (number of digits)	Maximum price to pay (number of digits)	Liquids
1	2	3	4	5	6	7	8
0.5	80	10	10	-40 to +50	99.99	9999.99	Petrol Diesel Kerosene

Important note: Apart from the mention of the certificates reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.