

OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R117/2007-NL1-16.05 revision 1 Project number 1901043

Page 1 of 3

Issuing authority Person responsible:

NMi Certin B.V. C. Oosterman

Applicant and Manufacturer

Endress + Hauser Flowtec AG

Kägenstrasse 7 CH-4153 Reinach Switzerland

Identification of the

certified type

A measurement transducer

Type: Promass Q 300 DNxxx[1]

Characteristics See page 2 and further

This Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML Type Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 117-1 (2007) "Dynamic measuring systems for liquids other than water"

Accuracy class 0,3 / 0,5 / 1,0 / 1,5

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Type Evaluation Report(s) is not permitted, although either may be reproduced in full.

[1] With xxx denoting the size of the Promass Q measurement sensor

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1

10 August 2017

C. Oosterman

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht the Netherlands T +31 78 6332332 certin@nmi.nl

www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org







OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R117/2007-NL1-16.05 revision 1 Project number 1901043 Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated report(s):

- NMi-15200323-01 dated 25 October 2016 that includes 94 pages
- No. NMi-16200475-01 dated 22 December 2016 that includes 247 pages;
- No. NMi-16200475-02 dated 22 December 2016 that includes 16 pages.

Characteristics of the flow transmitter

In Table 1 the general characteristics of the measuring instrument are presented. The construction of the measurement transducer is recorded in documentation folders TC7149-7 for the measurement sensor and TC10822-1 for the electronics.

Table 1 General characteristics of the Promass Q measurement sensor

+ + + + + + Sensor size	+ DN25 +	DN50	+ DN80	DN100
Maximum flow rate [t/h] + + +	+ -20+ -	+ 80 +	+ 200 +	+400 +
Minimum flow rate [t/h] + + +	0,45	+ 2 +	+ +6+ +	+ 14 +
Minimum Measured Quantity [kg]	+ +10+ +	+ 20 +	+ 100	+200 +

Maximum pressure + + + + 100 bar

Density range + + + + + + 400...1400 kg/m³.

Maximum viscosity 517 mPa·s
Temperature range ambient - 40...+55 °C

Product temperature range [°C] - 10...+90 °C for mass, density and volume measurement

-200...+90 °C for mass measurement

Accuracy Class 0.3; 0.5; 1.0 and 1.5

Environment class + + + + + M3 / E2 / H3

Table 2 General characteristics of the Promass 300 electronics

Environmental classes	M3/E2/H3
Ambient temperature range	-40+55 °C; condensing humidity
+ + + + + + + + + + + + + + + + + + +	24 VDC; 100240 VAC; 5060 Hz 24 VDC/100240 V AC;-/5060 Hz
Software identification	Version number: 01.00.02; CRC: 0xE87F (Modbus); 0x321F (Hart).
+ + + + + + + + + + + + + + + + + + + +	Version number: 01.00.03; * * * * * * * * * * * * * * * * * * *

The flow transmitter is solely to be used in combination of dynamic mass meters (Coriolis meters) of brand Endress + Hauser.



OIML Certificate of Conformity

OIML Member State The Netherlands

Number R117/2007-NL1-16.05 revision 1 Project number 1901043 Page 3 of 3

Certificate history:

This revision replaces the previous version.

Revision	Date	Description of the modification		
Initial +	22 December 2016			
+ + + + + + + + + + + +	7 August 2017	Maximum viscosity increased from 440 to 517 mPa·s. Based upon the measurements for the initial document on Aztec on the DN100 of 431,27 mPa·s increased with the 20% as indicated in OIML R117-1, § B.A.6.2.		