

OIML Member State
The Netherlands

Number R 117/2007-B-NL1-18.05 Revision 1
Project number 2655407
Page 1 of 3

Issuing authority
Person responsible: NMi Certin B.V.
M.Ph.D. Schmidt

Applicant and
Manufacturer: Advanced Flow Solutions d/b/a Liquid Controls
9201 N I-35 Service Rd,
Oklahoma City, OK 73131
United States of America

Identification of the
certified type: A **special gas extractor**
Type: EGS_1 and EGS_2

Characteristics: See page 2 and further

This OIML Certificate is issued under scheme B

This Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML Type Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 117-1 (2007) "Dynamic measuring systems for liquids other than water"

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Type Evaluation Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority: **NMi Certin B.V., OIML Issuing Authority NL1**
10 September 2021

Certification Board

OIML Member State
The Netherlands

Number R 117/2007-B-NL1-18.05 Revision 1
Project number 2655407
Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. CPC-9200487-01 dated 22 November 2010 that includes 14 pages;
- No. NMI-1901947-01 dated 26 November 2018 that includes 25 pages.

Characteristics of the special gas extractor

In Table 1 the general characteristics of the measuring instrument are presented.
The construction of the measuring instrument is recorded in the Documentation folder no. TC7329-5.

Part: Special gas extractor
Producer: Advanced Flow Solutions d/b/a Liquid Controls
Type: EGS_1 and EGS_2
Documentation folder: TC7329-5

Table 1 General characteristics of the gas elimination device type EGS_1 and EGS_2

There are two configurations of the special gas extractor:

	Type		Description
	EGS_1	EGS_2	
Gas eliminator ⁽¹⁾	421xx ⁽²⁾ ; A8xxx ⁽²⁾		
Filter type	F; FD ⁽³⁾ ; FMP	F	
Number of LC-sound gas indicators ⁽⁵⁾	2	1	
Two stage valve	Yes	Yes ⁽⁴⁾	
Intended for	Hydrocarbon based oils, fuels, chemical and industrial liquids with a viscosity of 1 – 5 mPa·s at 20 °C; LPG.		
Q _{max} [L/min]	See table 2 below.		
P(e) max [kPa]	1024; 2413 with F and FD filters 1930; 5100 with FMP filters. Equal to the maximum P(e) max. of the measuring system in which the Special Gas Extractor is used and not above 5100 kPa.		

Remarks:

- (1) The gas eliminator is mounted on top of the filter. Depending on the size, one or more gas eliminators are used.
- (2) Types 42100, 42101, 42104, A8170, A8180; A8200 and A8340A
- (3) FD filter is only used for the 4 inch diameter.
- (4) A two stage valve is used which is either pneumatically or hydraulically controlled.
For the EGS_2 a so called air check valve is required. See the documentation folder 7329-5.
- (5) The LC-sound gas indicator operates on the ultra sonic principle.

OIML Member State
The Netherlands

Number R 117/2007-B-NL1-18.05 Revision 1
Project number 2655407
Page 3 of 3

Table 2 Maximum flowrate of the gas elimination device type EGS_1 and EGS_2

Maximum Flowrate [L/min]	Test Characteristics	Remarks
500	2 LCS and 500 mm inlet pipe	Mounted on M7 meter
800	1 LCS and 500 mm inlet pipe	Mounted on M15 meter
1500	2 LCS and 500 mm inlet pipe	Mounted on M30 meter
3300	2 LCS	Mounted on MS75 meter

Production location

The special gas extractor is produced at one of the following production locations:

- Liquid Controls LLC, 105 Albrecht Drive, IL-60044, Lake Bluff, United States of America

Revision History

This revision replaces the previous versions.

Revision	Date	Change(s)
Initial	26 November 2018	Initial issue
1	10 September 2021	Company name changed