

OIML Member State
The Netherlands

Number R117/2019-A-NL1-21.06 revision 1
Project number 3639811
Page 1 of 6

Issuing authority NMI Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant Gilbarco Veeder Root
Burnt Mills Road, Basildon
Essex SS13 1DT
United Kingdom

Manufacturer Gilbarco Inc.
7300 W. Friendly Ave.
Greensboro, NC27420
United States of America

Identification of the certified type **A fuel dispenser**
Type: ENCORE

Characteristics See page 2 and further

This OIML Certificate is issued under scheme A

This Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML Type Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 117:2019 "Dynamic measuring systems for liquids other than water"

Accuracy class 0,5

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above.
This Certificate does not bestow any form of legal international approval.

This certificate and supporting reports comply with the requirements of OIML-CS-PD-07 clause 6.2.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Type Evaluation Report(s) is not permitted, although either may be reproduced in full. The conformity was established by the results of tests and examinations provided in the associated reports:

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**
15 August 2023

Certification Board

NMI Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



OIML Member State
The Netherlands

Number R117/2019-A-NL1-21.06 revision 1
Project number 3639811
Page 2 of 6

For evaluation of the complete fuel dispenser:

- No. NMI-2537261-02 dated 21 April 2021 that includes 59 pages.

For the mechanical meter sensors, brand Gilbarco Veeder-Root, type C+, V and V+ meter:

- No. CVN-10119469 dated 2 March 2001 that includes 56 pages;
- No. CVN-202211 dated 16 May 2003 that includes 49 pages;
- No. TR:1327 dated 15 April 2015 that includes 12 pages;
- No. TR:0561 dated 22 October 2009 that includes 14 pages;
- No. TR:0587 dated 29 September 2010 that includes 14 pages;
- No. TR:748 dated 10 May 2017 that includes 10 pages.

For the electronic calculating and indicating device, brand Gilbarco, type Sandpiper-III (USA):

- No. NMI-2537261-01 dated 21 April 2021 that includes 77 pages;
- No. NMI- 3639811-01 dated 15 August 2023 that includes 45 pages.

For the gas separator, brand Gilbarco Veeder-Root, type GPU90

- No. TR:740 dated 24 February 2017 that includes 11 pages;
- No. R117/1995-NL1-04.04 dated 24 January 2005 that includes 50 pages.

Characteristics of the fuel dispenser

In Table 1 the general characteristics of the measuring instrument are presented.

The construction of the measuring instrument is recorded under the type evaluation reports mentioned here above and the documentation folder no. NMI-2537261-02-1.

Table 1 General characteristics

| | |
|---------------------------------|--|
| Minimum – maximum flow rate | 1,6 – 40 L/min; Viscosity range 0,4 – 1,0 mPa·s. |
| | 2,0 – 80 L/min; Viscosity range 1,1 – 8,0 mPa·s. |
| Minimum measured quantity | 2, 5 and 10 L |
| Maximum pressure | 3,5 bar(g) |
| Accuracy class | 0,5 |
| Environmental classes | M1 / E1 |
| Ambient temperature range | -10 – +55 °C |
| Product temperature range | -10 – +50 °C |
| Intended for the measurement of | Hydrocarbon oils |

Each measuring system consists at least of:

- One combined pump and gas eliminator device (gas separator);
- If no gas separator (*) is used on the measurement system, provision shall be taken to prevent gas passing through the meter sensor.

OIML Member State
The Netherlands

Number R117/2019-A-NL1-21.06 revision 1
Project number 3639811
Page 3 of 6

- One or more meter sensor (meter);
- One calculating/indicating device (calculator).

The characteristics of the mentioned parts of the fuel dispenser are presented at table 2 and higher.

The same housing of the dispenser can comprise of one or more measuring systems. When more than one measuring systems are in one housing, one calculating/indicating device may be a common part of the measuring systems.

For multi-product dispensers it is only possible to deliver one product at the same time on one side of the dispenser.

The maximum flowrate once installed may be limited, but shall be greater than 10 times the minimum flowrate.

Table 2 gives an overview of flow characteristics of the configurations of the family of instruments.

Table 2 Flow characteristics of the configurations

| Configuration | Flow rate range | Remarks |
|---|-----------------|---|
| 1 x gas separator* 1 x meter sensor, type C+, V or V+ | 1,6 – 40 L/min | Intended for the measurement of hydrocarbon oils with viscosity range 0,4 – 1,0 mPa·s. |
| 1 x gas separator* 1 x meter sensor, type C+, V or V+ | 2,0 – 80 L/min | Intended for the measurement of hydrocarbon oils with viscosity range 1,1 – 8,0 mPa ·s. Optionally a feature to allow Qmax to be limited to 40 L/min |
| 1 x gas separator* 2 x meter sensors, type C+, V or V+, one per dispenser side. | 1,6 – 40 L/min | Intended for the measurement of hydrocarbon oils with viscosity range 0,4 – 1,0 mPa·s. The gas separator of this measuring system is suitable for use with two meter sensors. Each meter sensor is considered as part of an individual measuring system. Flowrate reduces to 40 L/min with both meter sensors operating. Optionally a feature to allow Qmax to be limited to 40 L/min with a single meter sensor operating. |

OIML Member State
The Netherlands

Number R117/2019-A-NL1-21.06 revision 1
Project number 3639811
Page 4 of 6

| Configuration | Flow rate range | Remarks |
|---|-----------------|--|
| 1 x gas separator* 2 x meter sensors, type C+, V or V+, one per dispenser side | 2,0 – 80 L/min | Intended for the measurement of hydrocarbon oils with viscosity range 1,1 – 8,0 mPa·s. The gas separator of this measuring system is suitable for use with two meter sensors. Each meter sensor is considered as part of an individual measuring system. Flowrate reduces to 40 L/min with both meter sensors operating. |

The configuration of the measuring system is recorded in the documentation folder number NMI-2583578-01-1

The complete family of dispensers consists of one family (which are of similar construction) and have the flow characteristics indicated in table 2.

Parts of the measuring system

The conformity of the following parts was established by the results of tests and examinations provided in the associated report(s):

Part: Measurement sensor
 Producer: Gilbarco Veeder Root
 Type: C+
 Reports: No. CVN-10119469 dated 2 March 2001 that includes 56 pages;
 No. CVN-202211 dated 16 May 2003 that includes 49 pages;
 No. TR:1327 dated 15 April 2015 that includes 12 pages;
 No.TR:748 dated 10 May 2017 that includes 10 pages.

Table 3 General characteristics of the measurement sensor type C+

| | | |
|---------------------------------|--|--|
| Flow rate range [L/min] | 1,6 – 40 L/min; | 2,0 – 80 L/min |
| Intended for the measurement of | Hydrocarbon oils with a viscosity of 0,4 mPa·s – 1,0 mPa·s | Hydrocarbon oils with a viscosity of 1,1 mPa·s – 8,0 mPa·s |
| MMQ | 1 L | 1 L |
| Maximum pressure | 3,5 bar | 3,5 bar |
| Environmental classes | M1 / E1 | M1 / E1 |
| Ambient temperature range | -10 °C / +55 °C | -10 °C / +55 °C |
| Product temperature range | -10 °C / +50 °C | -10 °C / +50 °C |

OIML Member State
The Netherlands

Number R117/2019-A-NL1-21.06 revision 1
Project number 3639811
Page 5 of 6

Part: Measurement sensor
 Producer: Gilbarco Veeder Root
 Type: V, V+
 Reports: No. TR:0561 dated 22 October 2009 that includes 14 pages;
 No. TR:0587 dated 29 September 2010 that includes 14 pages.

Table 4 General characteristics of the measurement sensor type V and V+

| | | |
|---------------------------------|--|--|
| Flow rate range [L/min] | 1,6 – 40 L/min | 2,0 – 80 L/min |
| Intended for the measurement of | Hydrocarbon oils with a viscosity of 0,4 mPa·s – 1,0 mPa·s | Hydrocarbon oils with a viscosity of 1,1 mPa·s – 8,0 mPa·s |
| MMQ | 2 L | 2 L |
| Maximum pressure | 3,5 bar | 3,5 bar |
| Environmental classes | M1 / E1 | M1 / E1 |
| Ambient temperature range | -10 °C / +55 °C | -10 °C / +55 °C |
| Product temperature range | -10 °C / +50 °C | -10 °C / +50 °C |

Part: Calculating/indicating device
 Producer: Gilbarco USA
 Type: Sandpiper-III
 Reports: No. NMI-2537261-01 dated 12 April 2021 that includes 80 pages.
 No. NMI-3639811-01 dated 15 August 2023 that includes 45 pages.

Table 5 General characteristics of the calculating/indicating device type Sandpiper-III (USA)

| | |
|---------------------------|--------------------------------------|
| Maximum volume indication | 7 digits (9999999; floating decimal) |
| Maximum unit price | 6 digits (999999; floating decimal) |
| Maximum price to pay | 7 digits (9999999; floating decimal) |
| Environmental classes | M1 / E1 |
| Ambient temperature range | -25 °C / +55 °C |
| Software identification | See table below. |
| Impulse encoder or pulser | TE CA – 1174 |

OIML Member State
The Netherlands

Number R117/2019-A-NL1-21.06 revision 1
Project number 3639811
Page 6 of 6

Table 6 Software versions and checksum of the calculating/indicating device type Sandpiper -III

| Software versions | CRC Checksum |
|--------------------------------|--------------|
| 04.0.66P (displayed as P04066) | DF5C |
| 04.1.27P (displayed as P04127) | 7E5A |
| 05.0.16P (displayed as P05016) | 0BB5 |

The software version is displayed during the power up sequence or through command code 1, whilst the checksum can be displayed through command code 2.

Part: Gas elimination device (gas separator)
 Producer: Gilbarco Veeder Root
 Type: GPU90
 Reports: No. TR:740 dated 24 February 2017 that includes 11 pages.
 NO. R117/1995-NL1-04.04 dated 24 January 2005 that includes 50 pages.

Table 7 General characteristics of the gas elimination device type GPU90

| | |
|---------------------------------|--|
| Maximum flow rate | 90 L/min |
| Minimum pressure | 1,4 bar |
| Maximum pressure | 3,0 bar |
| Environmental classes | M1 |
| Ambient temperature range | -40 °C / +55 °C |
| Product temperature range | -40 °C / +50 °C |
| Intended for the measurement of | low-viscosity mineral oils with a viscosity of 0,4 mPa·s – 8,0 mPa·s |

Certificate history:

This revision replaces the previous version.

| Revision | Date | Description of the modification |
|----------|----------------|---|
| Initial | 21 April 2021 | - |
| 01 | 15 August 2023 | Addition of new PCB and Display of the calculator |