

## **OIML** Certificate



Number R117/2019-A-NL1-23.01 revision 0 Project number 2451649 Page 1 of 5

Issuing authority Person responsible:

**OIML Member State** 

The Netherlands

NMi Certin B.V. M.Ph.D. Schmidt

Applicant and Manufacturer

KROHNE Ltd. 34 – 38 Rutherford Drive Park Farm Industrial Estate Wellingborough NN8 6AE United Kingdom

Identification of the certified type

A **measurement transducer** Type: OPTIMASS 2000; OPTIMASS 6000

Characteristics

See page 2 and further.

This OIML Certificate is issued under scheme A

This Certificate attests the conformity of the above-identified type (represented by the sample(s) identified in the OIML Type Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R 117-1:2019** "Dynamic measuring systems for liquids other than water"

Accuracy class 0.3 / 0.5

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Type Evaluation Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V.

The Netherlands

T +31 88 636 2332

Thijsseweg 11

2629 JA Delft

certin@nmi.nl

www.nmi.nl

NMi Certin B.V., OIML Issuing Authority NL1 13 January 2023

#### **Certification Board**

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.







**OIML Member State** 

The Netherlands

# OIML Certificate



Number R117/2019-A-NL1-23.01 revision 0 Project number 2451649 Page 2 of 5

The conformity was established by the results of tests and examinations provided in the associated reports:

Report number	Issue date	Number of pages				
Meas	surement sensor: OPTIMASS	2000				
NMi-2451649-05	13 January 2023	11				
Measurement sensor: OPTIMASS 6000						
NMi-245164901	24 November 2022	23				
	MFC400 electronics					
NMi-2451649-02	24 November 2022	46				

#### Characteristics of the measurement transducer

In tables 1 to 4, the general characteristics of the measuring instrument are presented. The construction of the measurement transducer is recorded in documentation folders TC11487-2 for the measurement sensor and TC11484-1 for the electronics.

### Table 1 General characteristics applicable to all OPTIMASS measurement sensors

685 ... 1100 kg/m<sup>3</sup>

- Density range:
- Accuracy class:
- 0.3 and 0.5 M3 / E3 / H3
- Environmental classes:Ambient temperature range:
- Ambient temperature range: -40 ... +55 °C
  Intended for the measurement of: Oil and oil products, chemicals and potable liquids,

Sensor Type <sup>(1)</sup>	Oil and oil products, chemicals, and potable liquids	Liquefied gases under pressure	Liquefied gases below -10 °C, cryogenic liquids, LNG, LCO <sub>2</sub>			
	Accuracy class					
	0.3; 0.5	1.0	1.5			
OPTIMASS 2000	Μ	-	-			
OPTIMASS 6000	MDV	-	-			

Notes:

(1) This table indicates the approved measurements: **M** for Mass, **D** for density, and **V** for volume.



**OIML Member State** The Netherlands



Number R117/2019-A-NL1-23.01 revision 0 Project number 2451649 Page 3 of 5

#### Table 2 Specific characteristics of the OPTIMASS 2000 measurement sensors

Senso	r size	DN100	DN150	DN250	DN400		
Maximum flow rate	[t/h]	220	500	1200	2985		
Minimum flow rate, class 0.3	[t/h]	11	25	60	150		
Minimum flow rate, class 0.5	[t/h]	6	16	40	100		
Minimum Measured Quantity	[kg]	200	200	500	1000		
Maximum viscosity [m	Pa∙s]	25	2	2	25		

### Further characteristics of the OPTIMASS 2000:

Accuracy Class	0.3	0.5	1.0	1.5
Maximum pressure	100 bar(g)		NA	NA
Temperature range liquid for mass measurement	-5 °C +85 °C		NA	NA
Temperature range liquid for density and volume measurement	NA		NA	NA

## Table 3 Specific characteristics of the OPTIMASS 6000 measurement sensors

Sense	or size	DN8	DN10	DN15	DN25	DN50	DN80
Maximum flow rate	[t/h]	0,6	1,2	3,8	19	35	80
Minimum flow rate, class 0.3	[t/h]	0,03	0,06	0,19	0,95	1,75	20
Minimum flow rate, class 0.5	[t/h]	0,015	0,03	0,095	0,475	0,875	10
Minimum Measured Quantity	[kg]	1	1	1	5	50	50
Maximum viscosity [r	nPa∙s]		(-	2	5		

Sens	or size	DN100	DN150	DN200	DN250	
Maximum flow rate	[t/h]	180	320	560	1000	
Minimum flow rate, class 0.3	[t/h]	28	70	148	270	
Minimum flow rate, class 0.5	[t/h]	14	35	74	135	
Minimum Measured Quantity	[kg]	200	100	100	200	
Maximum viscosity [	mPa∙s]			2	5	



**OIML Member State** The Netherlands



Number R117/2019-A-NL1-23.01 revision 0 Project number 2451649 Page 4 of 5

#### Further characteristics of the OPTIMASS 6000:

Accuracy Class	0.3	0.5	1.0	1.5
Maximum pressure	100 b	oar(g)	NA	NA
Temperature range liquid for mass measurement	-5 °C	+85 °C	NA	NA
Temperature range liquid for density and volume measurement	-5 °C	+85 °C	NA	NA

### Table 4 General characteristics of the MFC400 electronics

Environmental classes		M3 / E3 / H3	M3 / E3 / H3				
Ambient temperature range		-40+55 °C	-40+55 °C; condensing humidity				
Power supp	ply voltage		24 VDC 100240 V	24 VDC 100240 VAC, 5060 Hz			
Software id	entification						
ER version	Main software	User interface	Sensor electronics	SIL IO (exi)	IO2 software	Modbus	
ER 2.1.2_	V6.1.2_	V2.1.2_	V2.0.1_	V1.0.2_	V5.0.2_	V1.0.1_	
	0x01B6	0xBEF4CBA2	0xF9F6	0x1A2B	0xA6FE	0x353D6ABA	
ER 2.1.3_	V6.1.2_	V2.1.3_	V2.0.1_	V1.0.2_	V5.0.2_	V1.0.1_	
	0x01B6	0xAC61F43F	0xF9F6	0x1A2B	0xA6FE	0x353D6ABA	
ER 2.1.4_	V6.1.3_	V2.1.4_	V2.0.1_	V1.0.3_	V5.0.2_	V1.0.1_	
	0x71F9	0xB29DFE9A	0xF9F6	0x443D	0xA6FE	0x353D6ABA	
ER 2.1.5_	V6.1.3_	V2.1.4_	V2.0.1_	V2.0.0_	V5.0.2_	V1.0.1_	
	0x71F9	0xB29DFE9A	0xF9F6	0xA75B	0xA6FE	0x353D6ABA	
ER 2.1.6_	V6.1.4_	V2.1.6_	V2.0.1_	V2.0.0_	V5.0.2_	V1.2.0_	
	0x331C	0x75E94DAF	0xF9F6	0xA75B	0xA8FE	0x6289E485	

The ER number is directly linked to the hardware and the software of the instrument. If either one is updated, the number changes.

The MFC400 flow transmitter may only be used in combination with Coriolis measurement sensors manufactured by KROHNE Ltd.



# **OIML** Certificate

**OIML Member State** The Netherlands



Number R117/2019-A-NL1-23.01 revision 0 Project number 2451649 Page 5 of 5

## **Revision History**

Revision	Date	Change(s)	
0	13-01-2023	Initial issue	
	-		