



## OIML CERTIFICATE OF CONFORMITY

### Issuing Authority

Name: Centro Español de Metrología  
Address: C/ Alfar, 2 -28760 Tres Cantos - Madrid (Spain)  
Person responsible: José Ángel Terradillos. Head of Operations II division

### Applicant

Name: SERES ENVIRONNEMENT  
Address: 360 Rue Louis de Broglie, Parc de la Duranne-BP 20 087  
Aix en Provence (France)  
Manufacturer of the certified type is the applicant

### Identification of the certified type

Evidential breath analyzers SAF'IR Evolution

### Identification of the certified pattern:

Portable evidential breath analyzer without printer

<i>Model designation</i>	SAF'IR Evolution
Measurement range	0,00 mg/L to 2,00 mg/L
Scale interval	0,01 mg/L
Operating temperature range	0 °C to 40 °C
Warn up time	≤ 5 min

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report (s) with the requirements of the following Recommendation of the International Organization of Legal Metrology-OIML):

R 126 *Evidential breath analyzers* Edition 1998 (E)

This certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.


This certificate does not bestow any formal of legal international approval.



The conformity was established by the results of test and examinations provided in the associated Test Report:

N° - IT- 082365001 that includes 28 pages.

**The Issuing authority**



**Dr. Fernando Ferrer Margalef,**  
Director of CEM



**The OIML Member**



**Dr. Fernando Ferrer Margalef,**

Date: Tres Cantos, 6 November 2009

Additional characteristics

<i>Model designation</i>	<b>SAF'IR Evolution</b>
Storage temperature range	- 25 °C to 70 °C

The evidential breath analyzers contain an internal battery 12 VDC 9W max. The charger converts 100-240 V AC 50/60 Hz to 12 V DC 9 W

*Important note:* Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate is issued, partial quotation of the certificate and of the associated test report is not permitted, although either may be reproduced in full.