

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: SICK AG.
Address: Nimburger Strasse 11
D-79276 Reute
Germany


Manufacturer of the certified type

Name: SICK AG.
Address: Nimburger Strasse 11
D-79276 Reute
Germany

Identification of the certified type

Multi-dimensional measuring instrument for measuring cubic and rectangular, non-irregular shaped, non reflective and opaque boxes.
Type: VMS 510

	L	W	H
Max ≤	2000 mm	1000 mm	1000 mm
$V_{max} \leq 1.2 \text{ m/s}$			
Min ≥	50 mm	50 mm	50 mm
d ≥	5 mm	5 mm	5 mm
$V_{max} \leq 2 \text{ m/s}$			
Min ≥	100 mm	100 mm	50 mm
d ≥	10 mm	10 mm	5 mm



Nederlands Meetinstituut

OIML Member State
The Netherlands

OIML Certificate N° R129/2000-NL1-06.01

Project number 601851

Page 2 of 2

Maximum objects processed: 960 per minute

Temperature range 0 °C / + 40 °C

This Certificate attests the conformity of the above mentioned type (represented by the sample or samples identified in the associated Test Report, the description and the appertaining documentation folder), with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R129

Edition 2000 (E)

This Certificate relates only to the metrological and technical characteristics of the type of the instrument covered by the relevant OIML International Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

N° R129/2000-NL1-06.01A, that includes 64 pages;

N° R129/2000-NL1-06.01B, that includes 64 pages.

The Issuing Authority
Ing. C. Oosterman
Manager Product Certification

1/a

12 April 2006

*

**

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.