



OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: SICK AG.
Address: Nimburger Strasse 11
D-79276 Reute
Germany

Manufacturer of the certified type

Name: SICK AG.
Address: Nimburger Strasse 11
D-79276 Reute
Germany

Identification of the certified type

Multi-dimensional measuring instrument for measuring rectangular, non-rectangular, non-irregular shaped, non-reflective and opaque boxes.
Type: VMS 520

	L	W	H
Max ≤	2000 mm	1000 mm	1600 mm
Min ≥	50 mm	50 mm	50 mm
d ≥	5 mm	5 mm	5 mm
$V_{max} \leq 3 \text{ m/s}$			

Maximum objects processed: 2400 per minute

Temperature range -10 °C / + 40 °C



OIML Member State
The Netherlands

This Certificate attests the conformity of the above mentioned type (represented by the sample or samples identified in the associated Test Report, the EC-type-examination certificate and the description with number T10019 and the appertaining documentation folder), with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R129
Edition 2000 (E)

This Certificate relates only to the metrological and technical characteristics of the type of the instrument covered by the relevant OIML International Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Reports:
N° R129/2000-NL1-07.01, that includes 64 pages.

The Issuing Authority
Ing. C. Oosterman
Manager Product Certification



16 March 2007

*
* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.