



OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Accu-Sort Systems, Inc.
Address: 511 School House Road
Telford, PA 18969
United States of America

Manufacturer of the certified type

Name: Accu-Sort Systems, Inc.
Address: 511 School House Road
Telford, PA 18969
United States of America

Identification of certified type

Multi-dimensional measuring instrument
Type: DM-3500

	L	W	H
Max ≤	150 cm	120 cm	90 cm
Min ≥	10 cm	10 cm	5 cm
d ≤	0.5 cm	0.5 cm	0.5 cm



$v_{\min} \geq 0.2$ m/min
 $v_{\max} \leq 90$ m/min for Non-singulated objects
 $v_{\max} \leq 180$ m/min for Singulated objects

Temperature range $-10^{\circ}\text{C} / +40^{\circ}\text{C}$

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-examination certificate and the description with number T10088 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R129
Edition 2000 (E)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R129/2000-NL1-07.02A that includes 68 pages;
N° R129/2000-NL1-07.02B that includes 10 pages;
N° 700411 that includes 9 pages.

The Issuing Authority
Ing. C. Oosterman
Manager Product Certification


25 June 2007

*
*

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.