



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R129/2000-NL1-16.07
Project number 16200608
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Teraoaka Seiko Co., Ltd. 13-12 Kugahara, 5-Chome Ohta-Ku, Tokyo 146-8580 Japan
Identification of the certified type	A Multi-Dimensional Measuring instrument Type : SPK-2000 SPK-3000
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 129 - Edition 2000

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
5 April 2017


C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org



The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-16200608-04 dated 23 February 2017 that includes 70 pages;
- No. NMI-16200608-05 dated 23 February 2017 that includes 20 pages.

Characteristics of the multi-dimensional measuring instrument

Principle of operation	cutting a light beam		
SPK-2000			
Maximum dimension	Length	Width	Height
	max ≤ 1530 mm	max ≤ 600 mm	max ≤ 920 mm
Minimum dimension	min ≥ 150 mm	min ≥ 100 mm	min ≥ 60 mm
Scale interval d	d ≥ 5 mm		
SPK-3000I			
Maximum dimension	Length	Width	Height
	max ≤ 1530 mm	max ≤ 900 mm	max ≤ 920 mm
Minimum dimension	min ≥ 150 mm	min ≥ 100 mm	min ≥ 60 mm
Scale interval d	d ≥ 5 mm		
Measuring range	Single interval		
Speed range	30 m/min ≤ v ≤ 70 m/min		
Electromagnetic environment class	E2		
Mechanical environment class	M1		
Climatic environment	temperature range	5 °C / +40 °C	
	humidity	non-condensing	
	intended location	Closed	
Power supply voltage	3x 200 V AC 50/60 Hz		
Method of operation	automatic		
Limitations of use	rectangular, opaque, objects with regular surfaces		
Warm up time	8 minutes		
Minimum spacing between successive objects	spacing ≥ 70 mm		
Software identification	Scale (console)	Version number: 1.xx ^{*)}	
	Conveyor	Version number: 4.xx ^{*)}	

^{*)} xx is a number between 00 and 99