

Member State
Denmark

**OIML Certificate N°
R134/2003-DK1-09.01**

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: **The Danish Accreditation and Metrology Fund**
Address: Dyregårdsvej 5B
2740 Skovlunde
Denmark

Person responsible: P. Claudi Johansen

Applicant

Name: **Tunaylar Baskül San. ve. Tic. A.Ş**
Address: Beylikdüzü No: 6
P.K. 26
34520 Büyükçekmece - İstanbul
Turkey

Manufacturer

of the certified pattern: **Tunaylar Baskül San. ve. Tic. A.Ş**

Identification

of the certified pattern: **Automatic instrument for weighing road vehicles in motion**
Type: LL2/AW
Further characteristics are set out on page 2.

This certificate attests the conformity of the above mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R134 edition 2003,

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

OIML Certificate N° R134/2003-DK1-09.01

The conformity was established by tests described in the associated test reports from DELTA, DK, No. DANAK-1910523, dated 27 May. 2009, including 66 pages.

The issuing authority


The Danish Accreditation and Metrology Fund

Date: 11. JUN. 2009

The OIML member

P. Claus Johansen

Date: 11. JUN. 2009



Characteristics

Accuracy class		1	2
Maximum capacity	Max	30 000 kg	
Minimum capacity	Min	$\geq 50 \times d$	$\geq 10 \times d$
Scale interval	D	≥ 10 kg	50 kg
Maximum operation speed	v_{max}	3 km/h	5 km/h
Minimum operation speed	v_{min}	1 km/h	
Maximum transit speed:		30 km/h	
Maximum number of axles:		≤ 10	
Direction of travel		dual	
Scale interval for stationary load	E	$\geq 0.8 \mu V$	
Power supply voltage:		12 V _{DC}	
Connected load cells		Shall comply with R60	
Temperature range for the indicator		-10 °C / +40 °C	

Important note:

Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.