





OIML Certificate No. R137/2012-B-CN2-2018.02

OIML CERTIFICATE

ISSUED UNDER SCHEME B

OIML Issuing Authority

Name: National Institute of Metrology, China

Address: No.18, Bei San Huan Dong Lu, Chaoyang Dist, Beijing, P.R.China

Person responsible: Mr. Fang Xiang, Director

Applicant

Name: Qianwei Kromschroder Meters(Chongqing) Co., Ltd.

Address: Middle section No. 69, Huangshan Avenue, New North Zone,

Chongqing, P.R China

Manufacturer

Name: Qianwei Kromschroder Meters(Chongqing) Co., Ltd.

Address: Middle section No. 69, Huangshan Avenue, New North Zone,

Chongqing, P.R China

Identification of the

Diaphragm Gas Meter G6

certified type

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R137-1&2: 2012 (Including Amendment 2014)

R137-3: 2014

For accuracy class:

1.5

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

OIML Issuing Authority:

National Institute of Metrology, China

Date:

20 August 2018

No.18, Bei San Huan Dong Lu, Chaoyang Dist, Beijing, 100029, P.R.China +86 10 64524602 oimlia@nim.ac.cn www.nim.ac.cn

Fang Xiang Director

National Institute of Metrology, China

OIML Member State P. R. China

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. TER-R137/2012-CN2-2018.02 dated 15 August 2018 that includes 16 pages and the associated OIML test report:

No.

2018OM-002

dated

3 July 2018

that includes 44 pages

OIML Certificate History

Revision No.	Date	Description of the modification
0	20 August 2018	Initial issue

Characteristics of the Instrument

Manufacturer's trade mark	Qianwei Kromschroder Meters(Chongqing) Co., Ltd.
Type designation	G6
Accuracy class	1.5 class
Cyclic volume (if applicable)	2 dm³
Minimum pressure p_{\min}	0.2 kPa
Maximum pressure p_{max}	30 kPa
Ambient temperature range	-20 °C ∼ +50 °C
Gas temperature range	-20 °C ∼ +50 °C
Base pressure (if applicable)	N/A
Base temperature (if applicable)	N/A
$t_{\rm sp}$ (if applicable)	N/A
Electrical power	N/A
Identification of software	N/A