

OIML Member State
The Netherlands

Number R 137/2012-NL1-17.08
Project number 15200380
Page 1 of 2

Issuing authority
Person responsible: NMi Certin B.V.
C. Oosterman

Applicant and
Manufacturer Daesung Measuring CO., LTD.
662 kyungin-ro, Guro-gu, SungDong-Gu, Seoul, South Korea (152-888)

Identification of the
certified type **A diaphragm gas meter**
Type: DSM GxR

Characteristics See page 2 and further

This Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML Type Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 137-1 (2012) "Gas meters"

Accuracy class 1,5

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Type Evaluation Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
24 October 2017



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org



OIML Member State
The Netherlands

Number R 137/2012-NL1-17.08
Project number 15200380
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated report(s):

- No. NMI-15200380-02 dated 22 September 2017 that includes 44 pages.

Characteristics of the measuring instrument

In Table 1 the general characteristics of the measuring instrument are presented.
Table 2 gives an overview of the general characteristics of the family of instruments.
The construction of the measuring instrument is recorded in the Documentation folder no. T10708-1.

Table 1 General characteristics

Destined for the measurement of	Gas volume
Environmental classes	M1 / E1
Accuracy class	1,5
Maximum pressure	0,5 bar
Ambient temperature range	-10 – +55 °C
Gas temperature range	-10 – +55 °C
Designed for	Condensing humidity

Table 2 General characteristics of the family of instruments

Meter size	G1,6	G2,5	G4	G6
Minimum flow rate Q_{\min} (m ³ /h)	0,016	0,025	0,04	0,06
Transitional flow rate Q_t (m ³ /h)	0,25	0,4	0,6	1
Maximum flow rate Q_{\max} (m ³ /h)	2,5	4	6	10
Overload flow rate Q_r (m ³ /h)	3	4,8	7,2	12
Indicating range (m ³)	9999	99999	99999	99999
Verification scale interval (m ³)	0,0002	0,0002	0,002	0,002