

OIML Member State
The Netherlands

Number R137/2012-B-NL1-18.01
Project number 1901403
Page 1 of 4

Issuing authority NMI Certin B.V.
Person responsible: C. Oosterman

Applicant and Manufacturer MeterSit
Viale dell'Industria, 31
35129 Padova
Italy

Identification of the certified type **A Thermal Gas Meter**
Type: x485xxx

Characteristics See page 2 and further

This OIML Certificate is issued under scheme B

This Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML Type Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 137-1 (2012) "Gas meters"

Accuracy class 1.5

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Type Evaluation Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**
12 March 2018



C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org



OIML Member State
The Netherlands

Number R137/2012-B-NL1-18.01
Project number 1901403
Page 2 of 4

The conformity was established by the results of tests and examinations provided in the associated report(s):

- No. NMI-16200387-02 dated 17 October 2016 that includes 61 pages.
- No. NMI-16200852-01 dated 8 May 2017 that includes 13 pages;
- No. NMI-1901029-02 dated 11 October 2017 that includes 14 pages;
- No. NMI-1901403-02 dated 1 February 2018 that includes 21 pages.

Characteristics of the measuring instrument

In Table 1 the general characteristics of the measuring instrument are presented.
Table 2 gives an overview of the general characteristics of the family of instruments.
The construction of the measuring instrument is recorded in the Documentation folder no. T10362-16.

Table 1 General characteristics

Destined for the measurement of	Gas volume of natural gas, type H or L
Environmental classes	M1 / E2
Accuracy class	1.5
Maximum pressure	500 mbar
Ambient temperature range	-25 – +55 °C
Gas temperature range	-25 – +55 °C
Designed for	Condensing humidity
Orientation	Horizontal
Power supply voltage	Battery powered



OIML Certificate

OIML Member State
The Netherlands

Number R137/2012-B-NL1-18.01

Project number 1901403

Page 3 of 4

	Version number	Checksum	Meter size
	EL10	ADB2	G1.6
	EL10	ADB2	G2.5
Software identification	E132 E167 G182 G192 G193 I192 G194 GL01 GL10 EL10	03EF D029 A1A8 18FB 03B6 8F41 1CCF 5812 8096 ADB2	G4
	A132 A167 J182 J192 J193 L192 J194 JL01 JL10	CA53 7199 BDC1 3484 4586 D8DD 5FFA B0DE 7EEA	G6
	B166 B183 B192 B194 BL01	6CA4 82D8 B8EF 22FA BD57	G10
	F154 F166 C182 C192 C194 CL01	E336 7D4C C9BE BC94 F780 62F5	G16
	H154 H166 D182 D192 D194 DL01	6B95 F29E E589 E889 416D CBFE	G25



OIML Certificate

OIML Member State
The Netherlands

Number R137/2012-B-NL1-18.01

Project number 1901403

Page 4 of 4

Table 2 General characteristics of the family of instruments

Meter size	G1.6	G2.5	G4	G4 ext.	G6	G10	G16	G25
Minimum flow rate Q_{\min} (m ³ /h)	0,016	0,025	0,04	0,016	0,06	0,1	0,16	0,25
Transitional flow rate Q_t (m ³ /h)	0,25	0,4	0,6	0,25	1	1,6	2,5	4
Maximum flow rate Q_{\max} (m ³ /h)	2,5	4	6	6	10	16	25	40
Overload flow rate Q_r (m ³ /h)	3	4,8	7,2	7,2	12	19,2	30	48
Indicating range (m ³)	99999 or 999999			999999				
Verification scale interval (m ³)	0,001			0,001				
Nominal diameter [mm]	32			50		65		