



**OIML Member State**  
The Netherlands

# OIML Certificate

Number R 137/2012-B-NL1-19.13  
Project number 2403564  
Page 1 of 3

Issuing authority

NMi Certin B.V.  
Person responsible: F. van Booma-de Smit

Applicant

ZENNER Metering Technology (Shanghai) Ltd.  
NO.6558, East Yinggang Road  
Qingpu Industrial Zone, Shanghai  
P.R. China

Manufacturer

ZENNER Metering Technology  
(Shanghai) Ltd.  
NO.6558, East Yinggang Road  
Qingpu Industrial Zone, Shanghai  
P.R. China

ZENNER International GmbH & Co. KG  
Römerstadt 6 D  
66121 Saarbrücken  
Germany

Zenner do Brasil Instrumentos  
de Medição Ltda.  
Rua Batrolomeu de Gusmao  
2444-Novo Hamburgo-RS  
Brazil

ZENNER International GmbH & Co. KG  
Talstraße 2  
09619 Mulda  
Germany

ZENNER-COMA JVC.  
Construction Machinery Company  
125D Minh Khai  
Q Hai Ba Trung Hanoi  
Vietnam

ZENNER Aquamet India Pvt Ltd  
39-B HSIDC , Sec-31 Faridabad  
(Haryana)-121003  
INDIA

Zenner Performance Meters Inc.  
1910E. Westward Ave  
Banning, CA 92220  
United States of America

Issuing Authority

**NMi Certin B.V., OIML Issuing Authority NL1**  
18 November 2019

Certification Board

**NMi Certin B.V.**  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
[certin@nmi.nl](mailto:certin@nmi.nl)  
[www.nmi.nl](http://www.nmi.nl)

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



**OIML Member State**  
The Netherlands

Number R 137/2012-B-NL1-19.13  
Project number 2403564  
Page 2 of 3

Identification of the certified type

**Diaphragm gas meter**

Type: Atmos xxS  
(xx is G1.6, G2.5, G4, G6M, WG2.5 or WG6M)

Atmos HP xxA  
(xx is G1.6, G2.5, G4 or WG2.5)

Characteristics

See page 3

This OIML Certificate is issued under scheme B

This Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML Type Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R 137-1 (2012) "Gas meters"**

Accuracy class

1,5

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Type Evaluation Report(s) is not permitted, although either may be reproduced in full.

**OIML Member State**  
The Netherlands

Number R 137/2012-B-NL1-19.13  
Project number 2403564  
Page 3 of 3

The conformity was established by the results of tests and examinations provided in the associated report(s):

- No. NMI-13200090-04 dated 17 November 2015 that includes 50 pages;
- No. NMI-2403564-02 dated 18 November 2019 that includes 13 pages.

### Characteristics of the measuring instrument

In Table 1 the general characteristics of the measuring instrument are presented. Table 2 gives an overview of the general characteristics of the family of instruments.

**Table 1 General characteristics**

Destined for the measurement of	Gas volume
Environmental classes	M1 / E1
Accuracy class	1,5
Maximum pressure	Atmos xxS: 0,5 bar Atmos HP xxA: 1,5 bar
Ambient temperature range	-25 – +55 °C
Gas temperature range	-25 – +55 °C
Designed for	Condensing humidity
Orientation	Connection ports vertical

**Table 2 General characteristics of the family of instruments**

Meter size	G1.6	G2.5	G4	WG2.5	G4	G6M	WG6M
Minimum flow rate $Q_{\min}$ (m <sup>3</sup> /h)	0,016	0,025	0,04	0,016	0,04	0,06	0,04
Transitional flow rate $Q_t$ (m <sup>3</sup> /h)	0,25	0,4	0,6	0,25	0,6	0,1	0,6
Maximum flow rate $Q_{\max}$ (m <sup>3</sup> /h)	2,5	4	6	6	6	10	10
Overload flow rate $Q_r$ (m <sup>3</sup> /h)	3	4,8	7,2	7,2	7,2	12	12
Cyclic volume (dm <sup>3</sup> )	1,2				2,0		
Indicating range (m <sup>3</sup> )	99999				99999		
Verification scale interval (m <sup>3</sup> )	0,0002				0,0002		