



**OIML Member State**  
The Netherlands

# OIML Certificate

Number R 137/2012-B-NL1-19.14 revision 1  
Project number 2301193  
Page 1 of 3

Issuing authority NMI Certin B.V.  
Person responsible: F. van Booma-de Smit

Applicant and Manufacturer Tokyo Keiso Co., Ltd.  
Shiba Toho Bldg 1-7-24 Shibakoen  
105-8558 Minato-ku  
Japan

Identification of the certified type An **ultrasonic gas meter**  
Type: UTG1000

Characteristics See page 2 and further

This OIML Certificate is issued under scheme B

This Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML Type Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R 137-1 (2012) "Gas meters"**

Accuracy class 0,5

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Type Evaluation Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**  
30 December 2019

Certification Board

**NMI Certin B.V.**  
Thijssseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
[certin@nmi.nl](mailto:certin@nmi.nl)  
[www.nmi.nl](http://www.nmi.nl)

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated report:

- No. NMI-2301193-01 dated 20 December 2019 that includes 45 pages.

## Characteristics of the measuring instrument

In Table 1 the general characteristics of the measuring instrument are presented. Table 2 gives an overview of the general characteristics of the family of instruments.

**Table 1 General characteristics**

Destined for the measurement of	Gas volume
Environmental classes	M2 / E2
Accuracy class	0,5
Minimum and maximum pressure	4 ... 16 bar(a)
Gas, ambient and transducer temperature range	-40 ... +55 °C <sup>[1]</sup>
Designed for	Condensing humidity
Orientation	All orientations
Power supply voltage	20 – 30 V DC
Software identification	Version number: 1.0.2.3

Note [1]: the maximum temperature tests are performed at +60 °C. The instrument fully complied with the requirements at +60 °C. However, OIML R 137 section 5.1, limits this temperature to four specific values, in this case +55 °C.

**Table 2 General characteristics of the family of instruments<sup>[2]</sup>**

Diameter size		Maximum $Q_{max}$	Minimum $Q_t$	Minimum $Q_{min}$
Inch	DN	m <sup>3</sup> /h	m <sup>3</sup> /h	m <sup>3</sup> /h
6	150	2000	200	65
8	200	3600	360	120
10	250	5600	560	180

Note [2]: if higher values are chosen for  $Q_{min}$  or  $Q_t$  and/or lower values for  $Q_{max}$ , it has to be taken into account that the ratio  $Q_{max} / Q_{min} \geq 20$  and ratio  $Q_{max} / Q_t \geq 5$ .

## Installation conditions:

Any components which could affect the gas flow must be avoided within the prescribed inlet pipe length for mild and severe flow disturbances as specified below. The inlet pipe must be designed as a straight pipe section of the same nominal diameter as the gas meter. The meters shall be equipped with at least 5DN straight outlet pipe and the following installation conditions apply:

- Mild disturbance: 10D straight inlet pipe;
- Mild and severe disturbance: 20D straight inlet pipe;
- Mild and severe disturbance: 5D+FC+5D straight inlet pipe.

Flow conditioner (FC) type: Spearman. See type evaluation report NMI-2301193-01 for further details regarding the Spearman flow conditioner (drawing and dimensions).



**OIML Member State**  
The Netherlands

# OIML Certificate

Number R 137/2012-B-NL1-19.14 revision 1  
Project number 2301193  
Page 3 of 3

## Revision History

This revision replaces the previous version.

Revision	Date	Changes
Initial	20-12-2019	-
1	30-12-2019	Typing error in type name