



OIML Member State
The Netherlands

OIML Certificate

Number R137/2012-A-NL1-20.09 revision 1
Project number 2430358
Page 1 of 4

Issuing authority

NMi Certin B.V.
Person responsible: M. Boudewijns

Applicant

ZENNER Metering Technology (Shanghai) Ltd.
NO.6558, East Yinggang Road
Qingpu Industrial Zone, Shanghai
P.R. China

Manufacturer

ZENNER Metering Technology
(Shanghai) Ltd.
NO.6558, East Yinggang Road
Qingpu Industrial Zone, Shanghai
P.R. China

ZENNER International GmbH & Co. KG
Römerstadt 6 D
66121 Saarbrücken
Germany

Zenner do Brasil Instrumentos
de Medição Ltda.
Rua Batolomeu de Gusmao
2444-Novo Hamburgo-RS
Brazil

ZENNER International GmbH & Co. KG
Talstraße 2
09619 Mulda
Germany

ZENNER-COMA JVC.
Construction Machinery Company
125D Minh Khai
Q Hai Ba Trung Hanoi
Vietnam

ZENNER Aquamet India Pvt Ltd
39-B HSIDC , Sec-31 Faridabad
(Haryana)-121003
INDIA

Zenner Performance Meters Inc.
1910E. Westward Ave
Banning, CA 92220
United States of America

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
7 December 2020

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



OIML Member State
The Netherlands

Identification of the certified type

Diaphragm gas meter

Type: Atmos xxS
(xx is G1.6, G2.5, G4, G4L, G6M, WG2.5 or WG6M)

Atmos HP xxA
(xx is G1.6, G2.5, G4, G4L, G6M, WG2.5 or WG6M)

Characteristics See page 3

This OIML Certificate is issued under scheme A

This Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML Type Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R137-1 (2012) "Gas meters"

Accuracy class 1,5

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Type Evaluation Report(s) is not permitted, although either may be reproduced in full.

OIML Member State
The Netherlands

Number R137/2012-A-NL1-20.09 revision 1
Project number 2430358
Page 3 of 4

The conformity was established by the results of tests and examinations provided in the associated report(s):

- No. NMI-13200090-04 dated 17 November 2015 that includes 50 pages;
- No. NMI-2403564-02 dated 18 November 2019 that includes 13 pages;
- No. NMI-2449796-02A dated 9 June 2020 that includes 11 pages.
- No. NMI-2430358-02 dated 2 December 2020 that includes 18 pages.

This certificate and supporting reports comply with the requirements of OIML-CS-PD-07 clause 6.2.

Characteristics of the measuring instrument

In Table 1 the general characteristics of the measuring instrument are presented. Table 2 gives an overview of the general characteristics of the family of instruments.

Table 1 General characteristics

Destined for the measurement of	Gas volume
Environmental classes	M1 / E1
Accuracy class	1,5
Maximum pressure	Atmos xxS: 0,5 bar Atmos HP xxA: 1,5 bar
Ambient temperature range	-25 – +55 °C
Gas temperature range	-25 – +55 °C
Designed for	Condensing humidity
Orientation	Connection ports vertical

Table 2 General characteristics of the family of instruments

Meter size	G1.6	G2.5	G4	WG2.5	G4L	G6M	WG6M
Minimum flow rate Q_{\min} (m ³ /h)	0,016	0,025	0,04	0,016	0,04	0,06	0,04
Transitional flow rate Q_t (m ³ /h)	0,25	0,4	0,6	0,2	0,6	1	0,6
Maximum flow rate Q_{\max} (m ³ /h)	2,5	4	6	6	6	10	10
Overload flow rate Q_r (m ³ /h)	3	4,8	7,2	7,2	7,2	12	12
Cyclic volume (dm ³)	1,2				2,0		
Indicating range (m ³)	99999				99999		
Verification scale interval (m ³)	0,0002				0,0002		

OIML Member State
The Netherlands

Number R137/2012-A-NL1-20.09 revision 1
Project number 2430358
Page 4 of 4

Certificate history:

This revision replaces the previous version.

Revision	Date	Description of the modification
Initial	18 June 2020	-
1	7 December 2020	Added test report NMI-2430358-02 for single pipe and double pipe meter (steel) WG6M