



**OIML Member State**  
The Netherlands

# OIML Certificate

Number R137/2012-A-NL1-20.12 revision 3  
Project number 3569999  
Page 1 of 3

Issuing authority NMI Certin B.V.  
Person responsible: M.Ph.D. Schmidt

Applicant and Manufacturer Goldcard Smart Group Co., Ltd.  
No. 158 Jinqiao Street  
Qiantang District  
Hangzhou Zhejiang 310018  
P.R. China

Identification of the certified type **A diaphragm gas meter**  
Type: JGD1.6S-M, JGD2.5S-M, JGD4S-M

Characteristics See page 2 and further

This OIML Certificate is issued under scheme A

This Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML Type Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R 137-1 (2012) "Gas meters"**

Accuracy class 1.5

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above. This Certificate does not bestow any form of legal international approval.

This certificate and supporting reports comply with the requirements of OIML-CS-PD-07 clause 6.2.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Type Evaluation Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**  
24 February 2023

Certification Board

**NMI Certin B.V.**  
Thijssseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
[certin@nmi.nl](mailto:certin@nmi.nl)  
[www.nmi.nl](http://www.nmi.nl)

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-13200090-04 dated 17 November 2015 that includes 50 pages;
- No. NMI-2437830-01 dated 19 June 2020 that includes 41 pages;
- No. NMI-2502596-01 dated 16 April 2021 that includes 18 pages;
- No. NMI-2596227-01 dated 4 June 2021 that includes 16 pages.

### Characteristics of the measuring instrument

In Table 1 the general characteristics of the measuring instrument are presented. Table 2 gives an overview of the general characteristics of the family of instruments.

**Table 1 General characteristics**

|                                 |                           |
|---------------------------------|---------------------------|
| Destined for the measurement of | Gas volume                |
| Environmental classes           | M1 / E2                   |
| Accuracy class                  | 1,5                       |
| Maximum pressure                | 0,5 bar                   |
| Ambient temperature range       | -25 – +55 °C              |
| Gas temperature range           | -25 – +55 °C              |
| Designed for                    | Condensing humidity       |
| Orientation                     | Connection ports vertical |
| Power supply voltage            | 3,6 V DC                  |
| Software identification         | Version number: 01902506  |

**Table 2 General characteristics of the family of instruments**

| Meter size                                       | 1,6          | 2,5   | 2,5   | 4     | 4     | 4    |
|--|--------------|-------|-------|-------|-------|------|
| Minimum flow rate $Q_{\min}$ (m <sup>3</sup> /h) | 0,016        | 0,016 | 0,025 | 0,016 | 0,025 | 0,04 |
| Transitional flow rate $Q_t$ (m <sup>3</sup> /h) | 0,25         | 0,25  | 0,4   | 0,25  | 0,4   | 0,6  |
| Maximum flow rate $Q_{\max}$ (m <sup>3</sup> /h) | 2,5          | 4     | 4     | 6     | 6     | 6    |
| Overload flow rate $Q_r$ (m <sup>3</sup> /h)     | 3            | 4,8   | 4,8   | 7,2   | 7,2   | 7,2  |
| Indicating range (m <sup>3</sup> )               | 9999999,9999 |       |       |       |       |      |
| Verification scale interval (m <sup>3</sup> )    | 0,0001       |       |       |       |       |      |



**OIML Member State**  
The Netherlands

# OIML Certificate

Number R137/2012-A-NL1-20.12 revision 3  
Project number 3569999  
Page 3 of 3

## Certificate history:

This revision replaces the previous version.

| Revision | Date             | Description of the modification  |
|----------|------------------|--|
| Initial  | 13 July 2020     | -  |
| 1        | 19 April 2021    | Addition of models with Sigfox, GPRS and NB-IoT communication modules. |
| 2        | 7 June 2021      | Addition of NFC functionality.   |
| 3        | 24 February 2023 | Address change   |