



OIML Member State
P. R. China



OIML Certificate No.
R51/2006-A-CN2-23.01

OIML CERTIFICATE

ISSUED UNDER SCHEME A

OIML Issuing Authority Name: National Institute of Metrology, China
Address: No.18, Bei San Huan Dong Lu, Chaoyang Dist., Beijing, P.R.China
Person responsible: Mr. Fang Xiang, Director

Applicant Name: Bizerba (Shanghai) WeighTech & Systems Co., Ltd.
Address: D-3 No.195 Qianpu Road, East New Area of Songjiang Industrial Zone, Shanghai, P.R.China

Manufacturer Name: Bizerba (Shanghai) WeighTech & Systems Co., Ltd.
Address: D-3 No.195 Qianpu Road, East New Area of Songjiang Industrial Zone, Shanghai, P.R.China

Identification of the certified type Dynamic Checkweigher CWA 1500, CWA3000 (Further characteristics on page 2)

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R51 – Edition 2006

For accuracy class: XIII(1)

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

OIML Issuing Authority: **National Institute of Metrology, China**

Date: 07 April 2023



No.18, Bei San Huan Dong Lu,
Chaoyang Dist, Beijing,
100029, P.R.China
+86 10 64525646
oimlia@nim.ac.cn
www.nim.ac.cn

Fang Xiang
Director
National Institute of Metrology, China

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. TER-R51/2006-CN2-23.01 dated 06 April 2023 that includes 19 pages
and the associated OIML test report:

No. LSmm2023-01139 dated 28 March 2023 that includes 126 pages

OIML Certificate History

Revision No.	Date	Description of the modification
0	07 April 2023	Initial issue

Characteristics of the Instrument

Type	CWA1500	CWA3000
Variants	CWA1500	CWA3000
Accuracy class	XIII(1)	
Max	1500 g	3000 g
Min	20 g	40 g
<i>e</i>	0.5 g	1 g
<i>d</i>	0.5 g	1 g
<i>n</i>	3000	3000
Maxmim speed	50 m/min	50 m/min
Temperature range	0 °C ~ +40 °C	
Power supply	AC 110 V ~ 240 V, 50/60 Hz	
Software version	The control and display software version is emn.03.00, The weighing module software version is ECn.09.00	