



OIML Member State  
P. R. China



OIML Certificate No.  
R51/2006-A-CN2-24.01

## OIML CERTIFICATE

ISSUED UNDER SCHEME A

**OIML Issuing Authority** Name: National Institute of Metrology, China  
Address: No.18, Bei San Huan Dong Lu, Chaoyang Dist., Beijing, P.R.China  
Person responsible: Mr. Fang Xiang

**Applicant** Name: SENAD Technology Co., Ltd  
Address: 10th Floor, Building 5, No. 599, Huiwang East Road, Jiading District, Shanghai, P.R.China

**Manufacturer** Name: SENAD Technology Co., Ltd  
Address: 10th Floor, Building 5, No. 599, Huiwang East Road, Jiading District, Shanghai, P.R.China

**Identification of the certified type** Dynamic Scale SND-DWS-M (Further characteristics on page 2)

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R51 – Edition 2006**

For accuracy class: Y(a)

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

OIML Issuing Authority: **National Institute of Metrology, China**

Date: 05 June 2024

**Fang Xiang**

National Institute of Metrology, China

No.18, Bei San Huan Dong Lu,  
Chaoyang Dist, Beijing,  
100029, P.R.China  
+86 10 64525646  
oimlia@nim.ac.cn  
www.nim.ac.cn



The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. TER-R51/2006-CN2-24.01 dated 04 June 2024 that includes 19 pages

and the associated OIML test report:

No. LSmm2024-01512.1 dated 10 May 2024 that includes 41 pages

No. LSmm2024-01512.2 dated 10 May 2024 that includes 17 pages

### OIML Certificate History

Revision No.	Date	Description of the modification
0	05 June 2024	Initial issue

### Characteristics of the Instrument

Type	SND-DWS-M
Accuracy class	Y(a)
Max	60 kg
Min	0.3 kg
<i>e</i>	20 g
<i>d</i>	20 g
<i>n</i>	3000
Speed	90 m/min
Temperature range	5 °C - 40 °C
Power supply	AC 220 V, 50 Hz
Software version	FX5.70