



Physikalisch-Technische Bundesanstalt  
Braunschweig und Berlin

**OIML Member State**  
Germany

**OIML Certificate No.**  
**R51/2006-A-DE1-2023.01**

**OIML CERTIFICATE ISSUED UNDER SCHEME A**

**OIML Issuing Authority**

Name: Physikalisch-Technische Bundesanstalt,  
Conformity Assessment Body  
Address: Bundesallee 100, 38116 Braunschweig, GERMANY  
Person responsible: Dr.-Ing. Prof. h. c. Frank Härtig

**Applicant**

Name: Pfreundt GmbH  
Address: Robert-Bosch-Str. 5, 46354 Südlohn, Germany

**Manufacturer**

Name: Pfreundt GmbH  
Address: Robert-Bosch-Str. 5, 46354 Südlohn, Germany

**Identification of the certified type** *(the detailed characteristics will be defined in the additional pages)*

Automatic catchweigher  
Type: WK60

**Designation of the module** *(if applicable)*

Not applicable

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 51

Edition (year): 2006

For accuracy class (if applicable): Y(b)

**OIML Certificate No.  
R51/2006-A-DE1-2023.01**

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. 1.12-4108889 dated 07.03.2023 that includes 21 pages

The technical documentation relating to the identified type is contained in documentation file:

No. ZDS-R51/2006-A-DE1-2023.01 dated 07.03.2023 that includes 3 pages

**OIML Certificate History**

<b>Revision No.</b>	<b>Date</b>	<b>Description of the modification</b>
Initial issuance	07.03.2023	---

Identification, signature and stamp

**The Issuing Authority**

Dr. Oliver Mack

Member of Conformity Assessment Body

Date: 07.03.2023



Table 1: Essential Data

Model	1, 2, 3
Accuracy class	Y(b)
Maximum capacity Max	depending on the vehicle
Minimum capacity Min	$\geq 10 e$
Number n of scale intervals	$\leq 500$ $\leq 400$ <sup>a)</sup>
Verification scale interval	$\geq 2$ kg

<sup>a)</sup> with digital load cell CP9.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.