

OIML Member State

The Netherlands

OIML Certificate



Number R61/2017-A-NL1-22.01 revision 0 Project number 3269935 Page 1 of 4

Issuing authority NMi Certin B.V. Person responsible: M.Ph.D. Schmidt Applicant and PIM Products B.V. Manufacturer Herenweg 16 8536 TN Oosterzee The Netherlands

Identification of the certified type An Automatic gravimetric filling instrument Type : PIM MHWx (where x represents the number of load receptors)

Characteristics Se

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 61-1:2017 for reference class Ref (1)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority



NMi Certin B.V. Thijsseweg 11 2629 JA Delft The Netherlands T +31 88 6362332 certin@nmi.nl www.nmi.nl NMi Certin B.V., OIML Issuing Authority NL1 21 December 2022

Certification Board

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.







OIML Member State

The Netherlands

OIML Certificate



Number R61/2017-A-NL1-22.01 revision 0 Project number 3269935 Page 2 of 4

The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Reports:

- No. NMi-3269935-01 dated 21 December 2022 that includes 31 pages;
- No. NMi-3269935-02 dated 21 December 2022 that includes 19 pages;
- No. NMi-3269935-03 dated 21 December 2022 that includes 28 pages.

Characteristics of the automatic gravimetric filling instrument

Method of operation		selective combination weighing or selective combination weighing combined with cummulative filling			
		or			
		filling by one weighing cycle			
Reference accuracy class		Ref (1) the operational accuracy class X(x) is determined at the time of putting into use			
Electromagnetic enviror	nment class	E2			
	temperature range	-10 °C / +40 °C			
Climatic anvironment		ADPD and DDPD	Mechanical assembly		
	humidity	non-condensing	condensing		
Ū	intended location	Closed	open and closed		
Maximum capacity of ea	ach load receptor	Calculated using the compatibility of modules form, contained in OIML R76 (2006) clause F.4			
Scale interval		Calculated using the compatibility of modules form, contained in OIML R76 (2006) clause F.4			
Number of load recepto	rs	4 to 24			
Power supply voltage		24 V DC			
Software identification Monitor SW-version		V7,xxx (xxx is number between 001 and 999)			
Software identification BrainsBox SW-versie		zzzz-V7.xxx:yyyy-mm-dd (zzzz is number between 0000 and 9999) (xxx is number between 013 and 999) yyyy-mm-dd represents a date			

The identification number will be displayed on the Human Machine Interface (HMI) by:

- Selecting the button in the lower left corner of the main menu;
- Selecting the information button.
- (\bullet)



OIML Certificate

OIML Member State The Netherlands



Number R61/2017-A-NL1-22.01 revision 0 Project number 3269935 Page 3 of 4

Characteristics of the Analog Data Processing Device

Accuracy class	OIML R 76					
	OIML R 61	Ref(0,2)				
Weighing range	·	Single interval				
Maximum number	of scale intervals	n ≤ 3000				
Load cell excitation	n voltage	5 V AC square wave				
Minimum signal in	put voltage	U _{min} = 0 mV				
Minimum input vo interval	ltage per verification scale	1 μV				
Minimum load cell	resistance	350 Ω				
Maximum load cel	l resistance	1050 Ω				
Fraction of the ma	ximum permissible error	0,5				
Load cell interface		6-wire with sense technology, may be configured as 4-wire				
Maximum value of wire section betwe junction box or loa	f the cable length per cross een the indicator and the ad cells	No special cable length In case sense technology is not used the load cells are connected directly without junction box or extension cable				

Rated minimum fill (Minfill):

	Minfill [g]							
		Operational accuracy class X(1)						
Average number of loads per fill	1	2	3	4	+5	6	7	8
0,5 g	22	31,5	38,5	44,5	49,5	109	117,5	125,5
1 g	44	126	154	178	199	327	353	377
2 g	178	378	462	1066	1192	1306	1412	1508
5 g	1335	1885	2310	2665	2980	3265	3530	3770
10 g	2670	3770	4620	5330	5960	6530	7060	7540



OIML Member State The Netherlands

OIML Certificate



Number R61/2017-A-NL1-22.01 revision 0 Project number 3269935 Page 4 of 4

		Minfill [g]							
)	Average number of loads per fill	1	2	3	4	5	6	7	8
Ī	0,5 g	11	15,5	19	22	25	27	29,5	31,5
	1 g	22	31	38	44	50	109	118	126
	2 g	44	126	154	<mark>17</mark> 8	198	326	352	378
	5 g	335	470	1155	1335	1490	1635	1765	1885
	10 g	1330	1890	2310	2670	2980	3270	3530	3770

Revision History

Revision	Date	Change(s)
0	2022-12-21	Initial issue.

