



**Member State**  
**Switzerland**

**OIML Certificate No.**  
**R076/2006-A-CH1-2019.01 Rev. 5**

## OIML-CS CERTIFICATE ISSUED UNDER SCHEME A

### *Issuing authority*

*Name* **Federal Institute of Metrology METAS**  
Conformity Evaluation Body METAS-Cert

*Address* Lindenweg 50, 3003 Bern-Wabern, Switzerland

*Person responsible* Gulian Couvreur, Head of METAS-Cert

### *Applicant*

*Name* **Sartorius Lab Instruments GmbH & Co KG**

*Address* Otto-Brenner-Strasse 20, 37079 Göttingen, Germany

*Manufacturer* The manufacturer of the certified pattern is the applicant.

### *Identification of the certified type*

*Type* **Precision and analysis balances BC-E and BC-A Line**

*Module* **BC-EA to BC-EH, BC-EK, BC-AA to BC-AH**

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):



**OIML R 76-1, edition 2006**

for accuracy class(es) (I), (II)

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

## Annex to the OIML-CS Certificate No R076/2006-A-CH1-2019.01 Revision 5

### Metrological characteristics

| Type   | BC-EA<br>BC-AA    | BC-EB<br>BC-AB  | BC-EB<br>BC-AB  |
|--|-------------------|-----------------|-----------------|
| Accuracy Class                                 | I                 | I               | II              |
| Max  | 50 g - 320 g      | 50 g - 220 g    | 0,1 g - 100 g   |
| e  | 1 mg - 2 mg       | 1 mg - 2 mg     | 1 mg - 2 mg     |
| d  | 0.1 mg - 2 mg     | 0.1 mg - 2 mg   | 0.1 mg - 2 mg   |
| n  | ≤ 320 000         | ≤ 220 000       | ≤ 100 000       |
| Tare-balancing range                           | until 100% of Max |                 |                 |
| Temperature range 1                            | +17 °C / +27 °C   | +17 °C / +27 °C | +10 °C / +30 °C |
| Temperature range 2 <sup>1)</sup>              | +10 °C / +30 °C   | +10 °C / +30 °C | +10 °C / +30 °C |
| Nominal capacity of the load receptor          | 384 g             | 264 g           | 264 g           |
| Initial zero setting + dead load <sup>2)</sup> | ≤ 234 g           | ≤ 214 g         | ≤ 263 g         |
| Maximum weighing pan size                      | Ø 90 mm           | Ø 90 mm         | Ø 90 mm         |

| Type   | BC-EC<br>BC-AC    | BC-ED<br>BC-AD  |
|--|-------------------|-----------------|
| Accuracy Class                                 | I                 | II              |
| Max  | 500 g – 1 500 g   | 1 g - 650 g     |
| e  | 10 mg - 20 mg     | 0.01 g – 0.1 g  |
| d  | 1 mg - 20 mg      | 0.001 g – 0.1 g |
| n  | ≤ 150 000         | ≤ 65 000        |
| Tare-balancing range                           | until 100% of Max |                 |
| Temperature range 1                            | +17 °C / +27 °C   | +10 °C / +30 °C |
| Temperature range 2 <sup>1)</sup>              | +10 °C / +30 °C   | +10 °C / +30 °C |
| Nominal capacity of the load receptor          | 1 800 g           | 780 g           |
| Initial zero setting + dead load <sup>2)</sup> | ≤ 1300 g          | ≤ 779 g         |
| Maximum weighing pan size                      | Ø 120 mm          | Ø 120 mm        |

**Annex to the OIML-CS Certificate No R076/2006-A-CH1-2019.01 Revision 5**

| Type   | BC-EE<br>BC-AE    | BC-EF<br>BC-AF  |
|--|-------------------|-----------------|
| Accuracy Class                                 | (II)              |                 |
| Max  | 500 g - 6 200 g   | 500 g - 6 200 g |
| e  | 0.1 g - 1 g       | 0.1 g - 1 g     |
| d  | 0.01 g - 1 g      | 0.01 g - 1 g    |
| n  | ≤ 62 000          | ≤ 62 000        |
| Tare-balancing range                           | until 100% of Max |                 |
| Temperature range 1                            | +10 °C / +30 °C   |                 |
| Nominal capacity of the load receptor          | 7 440 g           | 7 440 g         |
| Initial zero setting + dead load <sup>2)</sup> | 6 940 g           | 6 940 g         |
| Maximum weighing pan size                      | 180 mm x 180 mm   | Ø 180 mm        |

| Type   | BC-EG<br>BC-AG     | BC-EH<br>BC-AH     |
|--|--------------------|--------------------|
| Accuracy Class                                 | (II)               |                    |
| Max  | 5 000 g - 12 200 g | 5 000 g - 12 200 g |
| e  | 1 g                | 1 g                |
| d  | 0.1 g - 1 g        | 0.1 g - 1 g        |
| n  | ≤ 12 200           | ≤ 12 200           |
| Tare-balancing range                           | until 100% of Max  |                    |
| Temperature range                              | +10 °C / +30 °C    |                    |
| Nominal capacity of the load receptor          | 14 640 g           | 14 640 g           |
| Initial zero setting + dead load <sup>2)</sup> | ≤ 9 640 g          | ≤ 9 640 g          |
| Maximum weighing pan size                      | 180 mm x 180 mm    | Ø 180 mm           |

**Annex to the OIML-CS Certificate No R076/2006-A-CH1-2019.01 Revision 5**

|  |                   |
|--|-------------------|
| Type   | BC-EK             |
| Accuracy Class                                 | Ⓛ                 |
| Max  | 50 g – 95 g       |
| e  | 1 mg              |
| d  | 0.01 mg - 1 mg    |
| n  | ≤ 95 000          |
| Tare-balancing range                           | until 100% of Max |
| Temperature range 1                            | +17 °C / +27 °C   |
| Temperature range 2 <sup>1)</sup>              | +10 °C / +30 °C   |
| Nominal capacity of the load receptor          | 114 g             |
| Initial zero setting + dead load <sup>2)</sup> | ≤ 64 g            |
| Maximum weighing pan size                      | Ø 90 mm           |

<sup>1)</sup> Only for weighing instruments with incorporated span adjustment device being automatically released.

<sup>2)</sup> The sum of Max, initial zero setting range and dead load shall not exceed the nominal load of the load receptor.

*The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation reports:*

| <b>No.</b>   | <b>Date</b> | <b>Including pages</b> |
|--------------|-------------|------------------------|
| 6030-1189-05 | 2022-06-16  | 21                     |

*The technical documentation relating to the identified type is contained in documentation file:*

| <b>Name</b>                         | <b>Date</b> | <b>Including pages</b> |
|-------------------------------------|-------------|------------------------|
| R076_2006-A-CH1-2019.01-05_LERD_V01 | 2022-06-16  | 8                      |

*OIML Certificate History:*

| <b>Revision No.</b> | <b>Date</b> | <b>Description of the modification</b>  |
|---------------------|-------------|---|
| 00                  | 2019-09-24  | First issue   |
| 01                  | 2020-09-21  | Extension to the types BC-EA to BC-EH, BC-AA to BC-AH                               |
| 02                  | 2020-11-03  | Editorial changes   |
| 03                  | 2021-04-15  | Rebranding, new design  |
| 04                  | 2021-08-12  | Extension to the type BC-EK, version of software                                    |
| 05                  | 2022-06-16  | Version of software added, reinforcement of power connector (only in documentation) |

**The OIML Issuing Authority CH1**

3003 Bern-Wabern, 2022-06-16

*Approved by*

Gulian Couvreur, Head of sector  
METAS-Cert



*Important note:*

Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full