



OIML Member State  
P. R. China



OIML Certificate No.  
R76/2006-A-CN2-22.03

## OIML CERTIFICATE

ISSUED UNDER SCHEME A

**OIML Issuing Authority** Name: National Institute of Metrology, China  
Address: No.18, Bei San Huan Dong Lu, Chaoyang Dist., Beijing, P.R.China  
Person responsible: Mr. Fang Xiang, Director

**Applicant** Name: JIANGSU XIANGKANG TECHNOLOGY CO., LTD  
Address: NO.2-5, building 7, Shenye Huafu garden, Zhonglou District,  
Changzhou City, Jiangsu, P.R.China

**Manufacturer** Name: JIANGSU XIANGKANG TECHNOLOGY CO., LTD  
Address: NO.2-5, building 7, Shenye Huafu garden, Zhonglou District,  
Changzhou City, Jiangsu, P.R.China

**Identification of the certified type** Electronic weighing scale D-TCS-YG (Further characteristics on page 2)

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R76 – Edition 2006**

For accuracy class: III

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

OIML Issuing Authority: **National Institute of Metrology, China**

Date: 26 September 2022

**Fang Xiang**  
Director  
National Institute of Metrology, China



No.18, Bei San Huan Dong Lu,  
Chaoyang Dist, Beijing,  
100029, P.R.China  
+86 10 64525646  
oimlia@nim.ac.cn  
www.nim.ac.cn

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. TER-R76/2006-CN2-22.03 dated 26 September 2022 that includes 25 pages  
and the associated OIML test report:

No. LSmm2022-02828 dated 17 August 2022 that includes 18 pages

### OIML Certificate History

Revision No.	Date	Description of the modification
0	26 September 2022	Initial issue

### Characteristics of the Instrument

Type	D-TCS-YG
Variants	D-TCS-YG(only one variant)
Accuracy class	III
Max	150 kg
Min	1 kg
<i>e</i>	50 g
<i>d</i>	50 g
<i>n</i>	3000
Tare-balancing range	-150 kg
Temperature range	-10°C ~ +40°C
Power supply	AC (100 ~ 240) V, (50/60) Hz
Software version	1.xx.xxxx (x = 0 ... 9)