



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-19.49 Revision 1
Project number 3567445
Page 1 of 2

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

**Applicant and
Manufacturer**

Changzhou Sunmi Scale Technology Co. Ltd.
Xinzha Technology Industrial Park
Building 19&21, 213000
Changzhou, Jiangsu
China

Identification of the
certified type

A Non-automatic weighing instrument
Type : S2

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class **III** or **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
9 December 2022

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org



The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Reports:

- No. NMI-2305042-01 dated 27 September 2019 that includes 63 pages;
- No. NMI-3567445-01 dated 9 December 2022 that includes 7 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III or IIII
Maximum capacity	$6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
Verification scale interval	$e \geq 2 \text{ g}$
Weighing ranges	Single interval Multi-interval
Maximum number of scale intervals	$n \leq 3000$
Maximum number of weighing ranges	2
Tare	$T \leq -(\frac{1}{2}\text{Max}-e)$ for instruments with one weighing range $T \leq -(\text{Max}_1-e_1)$ for multi-interval instruments
Temperature range	-10 °C / +40 °C
Power supply voltage	100 – 240 V AC 50/60 Hz
Application	Intended to be used for direct sales to the public
Software identification	Version number: 10xxx or 11xxx (x = 0... 9)

The software identification is displayed after selecting:

- "POS Steward" on the main screen;
- Pressing tag "Electronic scale";
- The software identification appears at the right bottom corner.

Revision History

This revision replaces the previous version.

Revision	Date	Change(s)
0	27 September 2019	Initial issue.
1	9 December 2022	New software version added.