



OIML Member State  
P. R. China



OIML Certificate No.  
R76/2006-A-CN2-23.01

## OIML CERTIFICATE

ISSUED UNDER SCHEME A

**OIML Issuing Authority** Name: National Institute of Metrology, China  
Address: No.18, Bei San Huan Dong Lu, Chaoyang Dist., Beijing, P.R.China  
Person responsible: Mr. Fang Xiang, Director

**Applicant** Name: Zhongshan Camry Electronic Co., Ltd.  
Address: Baishawan Industrial Park, Qiwan Road East, East District, Zhongshan, Guangdong, China

**Manufacturer** Name: Zhongshan Camry Electronic Co., Ltd.  
Address: Baishawan Industrial Park, Qiwan Road East, East District, Zhongshan, Guangdong, China

**Identification of the certified type** Price Computing Scale TCS series (Further characteristics on page 2)

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R76 – Edition 2006**

For accuracy class: **III**

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

OIML Issuing Authority: **National Institute of Metrology, China**

Date: 07 April 2023

**Fang Xiang**  
Director  
National Institute of Metrology, China



No.18, Bei San Huan Dong Lu,  
Chaoyang Dist, Beijing,  
100029, P.R.China  
+86 10 64525646  
oimlia@nim.ac.cn  
www.nim.ac.cn

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. TER-R76/2006-CN2-23.01 dated 06 April 2023 that includes 25 pages

and the associated OIML test report:

No. LSmm2023-00804.1 dated 07 March 2023 that includes 41 pages

No. LSmm2023-00804.2 dated 09 March 2023 that includes 15 pages

**OIML Certificate History**

Revision No.	Date	Description of the modification
0	07 April 2023	Initial issue

**Characteristics of the Instrument**

Variants of type	TCS-150-JE62Z TCS-150-JC62Z	TCS-300-JE62Z TCS-300-JC62Z
Accuracy class	III	
Maximum capacity, Max	150 kg	300 kg
Minimum capacity, Min	1 kg	2 kg
Verification scale interval, <i>e</i>	50 g	100 g
Actual scale interval, <i>d</i>	50 g	100 g
Verification scale intervals, <i>n</i>	3000	
Weighing range(s)	Single range	
Tare-balancing range	Until 100% of Max	
Temperature range	0 °C ~ 40 °C	
Power supply	AC (100~220) V, 50/60 Hz Rechargeable battery 4 V 4 Ah	
Software version	Vr1.69	