



Physikalisch-Technische Bundesanstalt
Braunschweig und Berlin
Nationales Metrologieinstitut

Member State of OIML
Germany



OIML Certificate No.
R76/2006-DE1-08.05
Revision 2

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Physikalisch-Technische Bundesanstalt
Address: Bundesallee 100, 38116 Braunschweig
Person responsible: Dr. O. Mack

Applicant

Name: Bizerba GmbH & Co. KG
Address: Wilhelm-Kraut-Str. 65, 72336 Balingen

Manufacturer of the certified type is the applicant.

Identification of the certified type

Non-automatic price-computing weighing instrument
for direct sales to the public
Type: K...

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, edition 2006,
for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.



OIML Certificate No.
R76/2006-DE1-08.05
Revision 2

This Revision 2 was issued because new variants of load-receptors, software updates and structural modifications were added. The conformity is established in the Report No. 1.12-4037754, Revision 2, (34 pages) and by tests described in the associated Test Reports

No. 1011040104	that includes 55 pages
No. 1104080104	that includes 15 pages
No. 1105070104	that includes 15 pages
No. 1203070104	that includes 16 pages
No. 1203010104	that includes 16 pages
No. 1205060104	that includes 7 pages
No. 1207020104	that includes 16 pages
No. 1301010104	that includes 17 pages
No. 1311010104	that includes 55 pages
No. 1408020104	that includes 17 pages
No. 1408030104	that includes 17 pages
No. 1408041104	that includes 17 pages
No. 1409010104	that includes 51 pages
No. 1505010104	that includes 17 pages

The test results of the former Test Reports No. 1.12-4037754/1 to 1.12-4037754/7 remain valid.

The Issuing Authority

Dr. O. Mack
Member of Certification Body

12.01.2016



The OIML Member

Dr. R. Schwartz
Vice President

12.01.2016

Identification of the pattern (continued)

The weighing instrument consists of a weighing module, comprising the load receptor with or without a lever system, the strain gauge load cell and the electronics with an AD-converter for signal processing, of an incorporated indicating device (Touch-Screen) for displaying the weighing results and to operate the instrument.

Table 1

Accuracy class	III
Type	K... “...” means optional alphanumeric string
Max ≤	≤ 30 kg, with build-in load-receptor ≤ 200 kg, with separate load-receptor
e =	1 g ... 50 g
d =	1 g ... 50 g
n ≤	7500
Tare balancing range (subtr.)	≤ 100 % of Max
Preset tare range	≤ Max ≤ Max ₁ at multi-interval instruments
Temperature range	-10 °C / 40 °C

The weighing ranges including Max, Min, e, d and number of verification scale intervals may be chosen within the limits of No. 3.2 of R76-1.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.