OIML REALINGATION SYSTEM	FORCE		
OIML Member State Denmark	OIML Certificate No. R76/2006-A-DK2-2023.03		
OIML CERTIFICATE ISSUED UNDER SCHEME A			
OIML Issuing Authority Name: FORCE Certification A/S Address: Park Allé 345, 2605 Brøndby, Denmark Person responsible: Per Rafn Crety			
Applicant Name: Xiamen ATP Technology Co., Ltd. Address: Room 816, Eastern Commercial Building Yibing North Road, Huli District Yiamen China China Manufacturer Xiamen ATP Technology Co., Ltd. Identification of the certified type (the detailed characteristics will be defined in the additional pages) ATP160xx (xx is indicating variants and may be A-Z and 0-99)			
Designation of the module (if applicable) Non-automatic weighing instrument			
This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):			
OIML R 76-1, Edition (year): 2006 For accuracy class (if applicable): III			

OIML Certificate No. R76/2006-A-DK2-2023.03

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML reports:

Type examination report: No. 121-33865.90.10, dated 10 August 2022, that includes 69 pages

Type evaluation report: No. 123-25950.90.20, dated 07 June 2023, that includes 22 pages

The technical documentation relating to the identified type is contained in documentation file:

121-33865

OIML Certificate History

Revision No.	Date	Description of the modification
Initial version	22 June 2023	
		/
Cer	Cation	SYSE
Identification, signature and The OIML Issuing Autho FORCE Certification A/S	l stamp	
Date: 22 June 2023		
Jens Hovgård Jensen		
Certification Manager		
Important note: Apart fro OIML M Certificat	ember State in which the Certific e and of the associated OIML type	s reference number and the name of the rate is issued, partial quotation of the pe evaluation report(s) is not permitted,
although	either may be reproduced in full.	

III

150

Descriptive annex

Accuracy class

Characteristics

•

•	Single interval, multi-interval (dual)		
•	Maximum number of verification scale intervals:	3000 or 2x3000	
•	Maximum capacity (Max):	6 kg to 30 kg	
•	Minimum capacity (Min):	$20 \times e$	
•	Verification scale interval(e):	$\geq 2 \text{ g}$	
•	Maximum tare effect:	- Max	
•	Temperature range.	-10 °C to +40 °C	
•	Power supply:	110/220 VAC 50/60Hz	
Software The approved so	oftware is version 1 BJ_1.0.1.2f		
This information is displayed at power up.			
	etting device ($\leq 20\%$ of Max) atic zero setting device ($\leq 4\%$ of Max)		

'Scatic

- -- Zero tracking device ($\leq 4\%$ of Max)
- Semi-automatic subtractive tare weighing device
- Preset Tare

- Label Printing
- Stable equilibrium, Zero and Net indicators.
- PLU tables
- Totalization _

Interfaces

- **RS232C** _
- Cash Drawer _
- Etherner _
- Wireless LAN _
- USB _

The interfaces do not have to be secured.