



OIML Member State
Denmark

OIML Certificate No.
R76/2006-A-DK2-24.02 Revision 1

OIML CERTIFICATE ISSUED UNDER SCHEME A

OIML Issuing Authority

Name: **FORCE Certification A/S**
Address: **Park Allé 345, 2605 Brøndby, Denmark**
Person responsible: **Per Rafn Crety**

Applicant

Name: **Suzhou Accurate Weigh Co., Ltd. / Suzhou Hope Scale Co., Ltd.**
Address: **4th Floor, No. 6 West Building, No. 1919 Fengyang Road,
Weitang Town, Xiangcheng District
Suzhou
China**

Manufacturer: Suzhou Accurate Weigh Co., Ltd. / Suzhou Hope Scale Co., Ltd.

Identification of the certified type (*the detailed characteristics will be defined in the additional pages*)

HTW0x / HTP0x / HTC0x / HTWSS0x

Designation of the module (*if applicable*)

Non-automatic weighing instrument

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1, Edition (year): 2006

For accuracy class (if applicable): **III**

**OIML Certificate No.
R76/2006-A-DK2-24.02 Revision 1**

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML reports:

Type examination report: No. 122-21294.10, dated 08 September 2022, that includes 66 pages

Type examination report: No. 123-33997.10, dated 26 February 2022, that includes 48 pages

Type evaluation report: No. 122-21294.90.20, dated 14 December 2022, that includes 24 pages

Type evaluation report: No. 123-33997.90.20, dated 17 April 2024, that includes 24 pages

The technical documentation relating to the identified type is contained in documentation file:
122-21294 and 123-33997

OIML Certificate History

Revision No.	Date	Description of the modification
Initial version	30 April 2024	-
1	23 September 2024	Year in certificate No. changed from 4 to 2 digits.

Identification, signature and stamp

The OIML Issuing Authority
FORCE Certification A/S

Date: 23 September 2024

Jens Hovgård Jensen
Certification Manager

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

Descriptive annex

Characteristics

- | | |
|---|--|
| • Accuracy class | III |
| • Single interval, multirange (dual) | |
| • Maximum number of verification scale intervals: | 3000 |
| • Maximum capacity | 3kg, 6kg, 15kg, 30 kg, 60kg |
| • Scale interval | Max/3000 |
| • Minimum capacity | 20*e |
| • Maximum tare effect: | -Max |
| • Circuit for remote sense: | present on the model with 7- |
| • Temperature range. | -10 °C to +40 °C |
| • Power supply: | 220VAC 50/60 Hz.
6 VDC rechargeable battery |

Software

The approved software is version VER-11
This information is displayed at power up.

Devices

- Initial zero setting device ($\leq 20\%$ of Max)
- Semi-automatic zero setting device ($\leq 4\%$ of Max)
- Zero tracking device ($\leq 4\%$ of Max)
- Semi-automatic subtractive tare weighing device
- Printing
- Stable equilibrium, Zero and Net indicators.
- Price computing
- Totalization
- Counting
- Option for battery operation

Interfaces

- RS-232C
- Ethernet
- RS485
- Bluetooth
- USB
- WiFi
- 4-20mA / 0-10V analogue output
- Digital output via Relay

The interface does not have to be secured.

