OIML RATION SUSPECTION	FORCE	
OIML Member State Denmark	OIML Certificate No. R76/2006-A-DK2-2024.07	
OIML CERTIFICATE ISSUED UNDER SCHEME A		
OIML Issuing Authority		
Name:FORCE Certification A/SAddress:Park Allé 345, 2605 Brøndby, DenmarkPerson responsible:Per Rafn Crety		
Applicant		
Name: Changzhou LINWT Instruments Co. Ltd.   Address: No.2, building 56, Tusincere Science Park,   No. 8 beitanghe Road, Tianing District Changzhou city   China China   Manufacturer: Changzhou LINWT Instruments Co. Ltd.   Identification of the certified type (the detailed characteristics will be defined in the additional pages)   ST510P		
<b>Designation of the module</b> ( <i>if applicable</i> )		
Non-automatic weighing instrument		
This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):		
OIML R 76-1, Edition (year): 2006		
For accuracy class (if applicable): III and IIII		

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML reports:

Type examination report: No. 123-30042.10, dated 17 January 2024, that includes 66 pages Type evaluation report: No. 123-30042.90.20, dated 20 June 2024, that includes 19 pages

The technical documentation relating to the identified type is contained in documentation file:

123-30042

## **OIML Certificate History**

Revision No.	Date	Description of the modification
Initial version	19 September 2024	/-
Certi	atio	SYSE
Identification, signature and s	stamp	
The OIML Issuing Authorit	•	
FORCE Certification A/S	•	
Date: 19 September 2024		
Jens Hovgård Jensen		
Certification Manager		
OIML Mer Certificate	nber State in which the Co	ficate's reference number and the name of the ertificate is issued, partial quotation of the IL type evaluation report(s) is not permitted, n full.

150

# **Descriptive annex**

#### Characteristics

Accuracy class:	III or IIII
Weighing range:	Single-interval,
Maximum number of Verification	
Scale Intervals:	$\leq 6000$
Maximum tare effect:	-Max
Fractional factor:	p'i = 0.5
Minimum input voltage per VSI:	≥1.67 μV
Excitation voltage:	5 VDC
Circuit for remote sense:	Yes, when using 6-wire connection
Minimum input impedance:	87 ohm
Maximum input impedance:	1100 ohm
Mains power supply:	24 VDC – not to be supplied from DC mains
Operational temperature:	-10 °C to +40 °C

### Software

The approved software version is v. 4.08 The version can be found in the menu.

### Devices

- Initial zero setting device ( $\leq 20\%$  of Max)
- Semi-automatic zero setting device ( $\leq 4\%$  of Max)
- Zero tracking device ( $\leq 4\%$  of Max)
- Semi-automatic subtractive tare weighing device
- Stable equilibrium, Zero and Net indicators.
- Option for battery operation

### Interfaces

- RS232 •
- RS485 •
- ProfiBus DP •
- Modbus RTU ٠
- Modbus-TCP ٠
- ProfiNet •
- CC-link ٠
- Ethernet/IP •
- 4-20 mA current loop •
- Digital input •
- Digital output •

The interface does not have to be secured.

Cation 3450