

		
<b>OIML Member State</b> Denmark	<b>OIML Certificate No.</b> R76/2006-A-DK2-2024.10 Revision 1	
<b>OIML CERTIFICATE ISSUED UNDER SCHEME A</b>		
<b>OIML Issuing Authority</b> Name: <b>FORCE Certification A/S</b> Address: <b>Park Allé 345, 2605 Brøndby, Denmark</b> Person responsible: <b>Per Rafn Crety</b>		
<b>Applicant</b> Name: <b>Stryker Medical</b> Address: <b>3800 E Centre St. Portage, Michigan 49002 USA</b>		
<b>Manufacturer</b> <b>Stryker Medical</b>		
<b>Identification of the certified type</b> <i>(the detailed characteristics will be defined in the additional pages)</i> <b>ProCuity MedSurg Bed</b>		
<b>Designation of the module</b> <i>(if applicable)</i> <b>Non-automatic weighing instrument</b>		
<p>This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):</p> <p><b>OIML R 76-1, Edition (year): 2006</b></p> <p>For accuracy class (if applicable): <b>III</b></p>		

**OIML Certificate No.**  
**R76/2006-A-DK2-2024.10 Revision 1**

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML reports:

Type examination report: No. 120-24467.10, dated 23 September 2020, that includes 65 pages

Type evaluation report: No. 124-29915.90.10, dated 05 August 2024, that includes 18 pages,

The technical documentation relating to the identified type is contained in documentation file:

120-24467

**OIML Certificate History**

<b>Revision No.</b>	<b>Date</b>	<b>Description of the modification</b>
Initial version	22 August 2024	-
Revision 1	02 September 2024	Correction of typo error

Identification, signature and stamp

**The OIML Issuing Authority**

FORCE Certification A/S

Date: 02 September 2024

Jens Hovgård Jensen

Certification Manager

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

## Descriptive annex

### Characteristics

Type:	ProCuity MedSurg Bed
Accuracy Class:	III
Weighing ranges:	Single-interval
Maximum capacity (Max):	250 kg
Minimum capacity (Min):	10 e
Verification scale interval (e=):	2 kg
Maximum tare effect:	- 60 kg
Operating temperature range:	5°C to 35°C
Power supply:	100 – 240 VAC 50/60 Hz

### Software

The legal software for the scale has its own version and checksum.  
These can be found in the setting menu.

Allowed version are

- v01.08.000.01 with checksum 0xE86A93DA
- v01.09.000.01 with checksum 0x77DAA10D
- v01.10.000.01 with checksum 0x22920465

### Devices

- Semi-automatic subtractive tare device
- Exchange equipment function<sup>\*)</sup>
- Weight logging device
- Operator information messages
- Gravity compensation device (sealed)
- Stable equilibrium and Net indicators.

<sup>\*)</sup> Lock the present net weight and take any change in the gross weight as a change in tare weight.

### Interfaces

The scale has no interfaces.