



| | | |
|---|---|--|
|  |  |  |
| OIML Member State Denmark | OIML Certificate No. R76/2006-A-DK2-24.12 | |
| OIML CERTIFICATE ISSUED UNDER SCHEME A | | |
| OIML Issuing Authority Name: FORCE Certification A/S Address: Park Allé 345, 2605 Brøndby, Denmark Person responsible: Per Rafn Crety | | |
| Applicant Name: Curiotec Co. Ltd., Address: 79, Myeong-bong-san-ro 352 beon-gil, guangton-mueon, Paju-si, Gyeonggi-do, 413-855, South Korea | | |
| Manufacturer Curiotec Co. Ltd., CAS (Zhejiang) Electronics Co. Ltd, China. CAS Corporation, Republic of Korea CAS Elektronik San. Tic. A.S., Turkey CAS Deutschland AG, Germany. | | |
| Identification of the certified type <i>(the detailed characteristics will be defined in the additional pages)</i> CTI-1200 series | | |
| Designation of the module <i>(if applicable)</i> Non-automatic weighing instrument | | |
| This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML): OIML R 76-1, Edition (year): 2006 For accuracy class (if applicable): III | | |

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML reports:

Type examination report:

TR-572, dated 11 May 2010, that includes 39 pages

No. SN1135, dated 11 May 2010, that includes 13 pages

No. SN1136, dated 11 May 2010, that includes 10 pages

No. SN1349, dated 11 March 2016, that includes 09 pages

No. SN1406, dated 18 October 2017, that includes 10 pages

Type evaluation report: No. 124-31193.90.10, dated 29 November 2024, that includes 17 pages

The technical documentation relating to the identified type is contained in documentation file:
124-31193.10

OIML Certificate History

| Revision No. | Date | Description of the modification |
|---------------------|------------------|--|
| Initial version | 12 December 2024 | |
| | | |
| | | |
| | | |

Identification, signature and stamp

The OIML Issuing Authority

FORCE Certification A/S

Date: 12 December 2024

Jens Hovgård Jensen

Certification Manager

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

Descriptive annex

Characteristics

The main features of the instruments are,

- Plastic construction (CTI-1200A and CTI-1201A)
- Metal construction (CTI-1200S and CTI-1200SC)
- Alphanumeric keypad
- LCD display and indicators (CTI-1201A)
- LED display and indicators (CTI-1200A, CTI-1200S and CTI-1200SC)
- Battery indicator

Software

The software is designated “V1.xx” or “V2.xx”

This information is displayed at power up and may be displayed like this:

where:

- xx is reflecting non-legally relevant changes and may be numbers, letters, symbols or blank,
(Note V can be displayed as v or V on an LED display)

Access to the legally relevant parameters is only possible by accessing the calibration switch on the main board. Access to this calibration switch and download of software is prevented by sealing the enclosure.

Examples of scales

| Designation | Platform | Dead load of receptor | Max (kg) | E= (kg) | Load cell type | Load cell E _{max} | Number of load cells |
|----------------------|--------------|-----------------------|----------|---------|----------------|----------------------------|----------------------|
| Indicator CTI-1200x | | | | | | | |
| CTI1200x-6 | SPS(SUS)-6 | 3 | 6 | 0,002 | BCLS-10L | 10 | 1 |
| CTI1200x-15 | SPS(SUS)-15 | 4 | 15 | 0,005 | BCLS-20L | 20 | 1 |
| CTI1200x-30 | SPS(SUS)-30 | 4 | 30 | 0,01 | BCLS-50L | 50 | 1 |
| CTI1200x-60 | SPS(SUS)-60 | 10 | 60 | 0,02 | BCLS-100L | 100 | 1 |
| CTI1200x-150 | SPS(SUS)-150 | 10 | 150 | 0,05 | BCLS-180L | 180 | 1 |
| Indicator CTI-1200Sx | | | | | | | |
| CTI1200Sx-6 | SPS(SUS)-6 | 3 | 6 | 0,002 | BCLS-10L | 10 | 1 |
| CTI1200Sx-15 | SPS(SUS)-15 | 4 | 15 | 0,005 | BCLS-20L | 20 | 1 |
| CTI1200Sx-30 | SPS(SUS)-30 | 4 | 30 | 0,01 | BCLS-50L | 50 | 1 |
| CTI1200Sx-60 | SPS(SUS)-60 | 10 | 60 | 0,02 | BCLS-100L | 100 | 1 |
| CTI1200Sx-150 | SPS(SUS)-150 | 10 | 150 | 0,05 | BCLS-180L | 180 | 1 |

*) E_{max} in the above table refers to the actual measuring range and does not include the dead load for the instrument

The load cell fitted in the instrument is a CAS load cell, according to the tables in section 3.1.1.

Technical data

| | |
|--|--|
| Power Supply | 12 VDC from external mains adapter intended for 100-240VAC 50/60Hz |
| Maximum number of scale intervals | 10,000 |
| Load cell excitation voltage | 5 VDC |
| Minimum load cell impedance | 43.75 Ω |
| Maximum load cell impedance | 1000 Ω |
| Minimum input voltage per verification scale interval | 0.5 μV |
| Measuring range minimum voltage | 0 mV |
| Measuring range maximum voltage | 16 mV |
| Fraction of maximum permissible error | P _{ind} =0.5 |
| Operating temperature | -10°C to -40°C |
| Load cell cable maximum length (From indicator to load cell junction box) | 22m/mm ² (6-wire configuration) |

Devices

- Initial zero setting device ($\leq 20\%$ of Max)
- Semi-automatic zero setting device ($\leq 4\%$ of Max)
- Zero tracking device ($\leq 4\%$ of Max)
- Zero indicator
- Net indicator
- Semi-automatic subtractive tare balancing device
- Gravity compensation
- Printing
- Counting function (CTI-1200A)
- Hold function
- Percent function (CTI-1200A)
- Totalisation (CTI-1200A)
- Checkweighing (CTI-1200A and CTI-1200SC)

Interfaces

RS232 / RS485.
USB

