

	
<b>OIML Member State</b> Japan	<b>OIML Certificate No.</b> R76/2006-A-JP1-25.01
<b>OIML CERTIFICATE ISSUED UNDER SCHEME A</b>	
<b>OIML Issuing Authority</b> Name: National Metrology Institute of Japan /National Institute of Advanced Industrial Science and Technology (NMIJ/AIST) Address: AIST Tsukuba Central 3, 1-1-1 Umezono Tsukuba Ibaraki 305-8563, Japan Person responsible: ISHIMURA Kazuhiko, President of AIST	
<b>Applicant</b> Name: InBody Co., Ltd. Address: 15, Heugam-gil, Ipjang-myeon, Seobuk-gu Cheonan-si Chungcheongnam-do 31025 Korea	
<b>Manufacturer</b> Name: InBody Co., Ltd. Address: 15, Heugam-gil, Ipjang-myeon, Seobuk-gu Cheonan-si Chungcheongnam-do 31025 Korea	
<b>Identification of the certified type</b> (the detailed characteristics will be defined in the additional pages) Models: InBody770S	
<b>Designation of the module</b> (if applicable) Non-automatic weighing instruments	
<p>This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):</p> <p><b>OIML R 76-1, Edition: 2006</b></p> <p>For accuracy class: III</p>	

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. 2024-003, dated 28 March 2025, that includes 5 pages

The technical documentation relating to the identified type is contained in documentation file:

No. 2024-003-D, dated 28 March 2025

#### OIML Certificate History

Revision No.	Date	Description of the modification
Revision 0	31 March 2025	OIML Certificate first issued

This revision replaces previous versions of the certificate.

Identification, signature and stamp

**The Issuing Authority**

NMIJ/AIST

**The OIML Member**

ISHIMURA Kazuhiko  
President of AIST  
31 March 2025



OTA Akihiro  
31 March 2025

大田 明博

The accreditation body:

NMIJ/AIST has achieved accreditation under the ASNITE-Product (OIML) scheme of IAJapan, which applies ISO/IEC 17065:2012 and regulations relevant to OIML-CS as the accreditation criteria. The accreditation identification for this accreditation is ASNITE 0001 Product and the details of the accreditation information could be referred from the IAJapan website (<https://www.nite.go.jp/en/iajapan/asnite/lab/index.html>).

Important note:

Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

## DESCRIPTIVE ANNEX

### Characteristics of the instrument:

The InBody770S is a class III, self-indicating, non-automatic weighing instrument.  
The instruments are not designed for direct sales to the public.

### Technical data:

Type	InBody770S
Class	III
Max	270 kg
e	0.1 kg
Min	2 kg
Temperature range	+ 10 °C to + 40 °C

### Device:

- Initial zero-setting device ( $\pm 10.1$  kg)
- Semi-automatic zero-setting device ( $\leq 4\%$  of Max)
- Zero-tracking ( $\leq 4\%$  of Max)
- Preset tare device ( $\leq 5$  kg)

### Interfaces:

- RS232C
- USB Host
- USB Slave
- Ethernet
- Wireless communication (Wi-Fi and Bluetooth)

### Software:

The legally relevant software is designated version IB770S-1.x, with x reflecting non-legally relevant changes.

### Sealing:

The enclosure of the weighing part is sealed with void seals.