

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-15.15 revision 2  
Project number 3510807  
Page 1 of 2

Issuing authority

NMi Certin B.V.  
Person responsible: M.Ph.D. Schmidt

Applicant and  
Manufacturer

TANITA Corporation  
14-2, 1-chome, Maeno-cho, Itabashi-ku  
Tokyo 174-8630  
Japan

Identification of the  
certified type

**A Non-automatic weighing instrument / Indicator**

Type : DC-430x

'x' denotes a placeholder for alternative variants not affecting  
metrological parameters

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76-1:2006** for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.

This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

**NMi Certin B.V., OIML Issuing Authority NL1**  
15 September 2022

Certification Board

NMi Certin B.V.  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-15.15 revision 2  
Project number 3510807  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Reports:

- No. NMI-14200589-01 dated 9 March 2015 that includes 23 pages;
- No. NMI-14200589-02 dated 9 March 2015 that includes 11 pages;
- No. NMI-12200474-01 dated 23 January 2013 that includes 31 pages;
- No. NMI-12200160-01 revision 1 dated 23 January 2013 that includes 54 pages;
- No. NMI-3510807-01 dated 13 September 2022 that includes 10 pages;
- No. NMI-3510807-02 dated 13 September 2022 that includes 14 pages.

### Characteristics of the non-automatic weighing instrument:

|                                   |   |
|-----------------------------------|---|
| Accuracy class                    | III   |
| Maximum capacity                  | Max $\leq$ 270 kg   |
| Verification scale interval       | $e \geq 0,1$ kg   |
| Weighing range                    | Single interval   |
| Maximum number of scale intervals | $n \leq 2700$ divisions   |
| Temperature range                 | +5 °C / +35 °C  |
| Power supply voltage              | 12 V DC supplied by a 100-240 V AC 50/60 Hz switching power adapter |
| Application                       | Intended to be used for weighing patients                           |
| Software identification           | Display part: Dc430.9321<br>Scale part: Wb260.9201                  |

### Revision History

This revision replaces the previous version.

| Revision | Date       | Change(s)   |
|----------|------------|---|
| 0        | 2015-03-13 | Initial issue.  |
| 1        | 2022-09-13 | Adding test reports for alternative adapter and changing the type name to DC-430x |
| 2        | 2022-09-15 | Correction of power supply description  |