

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-20.18 revision 5  
Project number 3741825  
Page 1 of 3

Issuing authority

NMi Certin B.V.  
Person responsible: M.Ph.D. Schmidt

Applicant and  
Manufacturer

Bizerba SE & Co. KG  
Wilhelm-Kraut-Straße 65  
72336 Balingen  
Germany

Identification of the  
certified type

**A Non-automatic weighing instrument**  
Type

: K3...

'...' denotes a placeholder for  
alternative variants not affecting  
metrological parameters.

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76-1:2006** for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

**NMi Certin B.V., OIML Issuing Authority NL1**  
29 August 2024

Certification Board

NMi Certin B.V.  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated OIML Reports:

- No. NMI-2243314-01 revision 1 dated 12 August 2019 that includes 55 pages;
- No. NMI-2409818-01 dated 18 February 2020 that includes 11 pages;
- No. NMI-2431931-01 dated 14 April 2020 that includes 45 pages;
- No. NMI-2477454-01 dated 15 April 2021 that includes 51 pages;
- No. NMI-2477454-02 dated 15 April 2021 that includes 18 pages;
- No. NMI-2529842-01 dated 15 April 2021 that includes 17 pages;
- No. 1.12-4037754/6 dated 30 May 2011 that includes 56 pages;
- No. NMI-3001387-01 dated 18 October 2022 that includes 26 pages;
- No. NMI-3001387-02 dated 18 October 2022 that includes 14 pages;
- No. NMI-3741825-01 dated 29 August 2024 that includes 15 pages.

### Characteristics of the non-automatic weighing instrument:

Accuracy class		III
Maximum capacity	Weighing module K3	$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
	Weighing module ScalePad USB	$\text{Max} \leq 15 \text{ kg}$
	Weighing module ADC-Box USB	Calculated using the compatibility of modules form, contained in EN 45501:2015 clause F.4
	Weighing module QBase	$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
Verification scale interval		$e \geq 1 \text{ g}$
Weighing range(s)		Single interval Multi-interval
Maximum number of scale intervals (one weighing range)		$n \leq 6000$
Maximum number of scale intervals (multi-interval)		$n \leq 3000$ (per partial weighing range)
Maximum number of partial weighing ranges		3
Semi-automatic Tare		$T \leq -100\% \text{ of Max}$
Preset Tare		100% of Max; 100% of $\text{Max}_1$ , in case of multi-interval instruments
Temperature range		$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Power supply voltage		120 – 240 V AC 50/60 Hz
Application		Intended to be used for direct sales to the public

Terminal software identification:

	Software version number (legally relevant) <sup>1)</sup>	Software ID (Windows)	Software ID (Linux)
Scale OEM Module	013	7640	0369
	015	8282	1171

A/D board software identification:

	Version number (legally relevant) <sup>1)</sup>	Software ID
ADW	010	6153
	011	5232
	030	1446
	014	4751

- <sup>1)</sup> The identification of the non-legally relevant software version can optionally be added to the identification of the legally relevant software version, separated by "::" (example: "013::x.y" or "010::x.y" or "030::x.y" or "014::x,y").

## Revision History

This revision replaces the previous versions.

Revision	Date	Changes
0	2020-04-14	-
1	2020-07-20	Correction of the maximum capacity specification (30 kg instead of 15 kg).
2	2021-04-15	Added Scale OEM Module software version 015 and ID's 8282 and 1171 Added OIML report NMI-2477454-01
3	2022-10-18	Adding ADPD ADC-Box USB and weighing modules ScalePad USB and QBase
4	2023-01-05	Editorial changes to the table of characteristics. Addition of a new software version.
5	2024-08-29	Added a new software version Added report NMI-3741825-01