OIML Certificate

Issuing authority	NMi Certin B.V. Person responsible: M.Ph.D. Schmidt
Applicant and Manufacturer	Mettler Toledo GmbH Im Langacher 44 8606 Greifensee Switzerland
Identification of the certified type	An **Indicator / Terminal** Type : IND890-2, IND970
Characteristics	See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class II or III or IIII

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority   **NMi Certin B.V., OIML Issuing Authority NL1**
18 February 2025

Certification Board

OIML Certificate

The conformity was established by the results of tests and examinations provided in the associated reports:

Indicators:
- No. NMi-1900621-02 dated 24 May 2017 that includes 14 pages;
- No. NMi-1901844-01 dated 28 May 2018 that includes 19 pages;
- No. NMi-2657900-01 dated 12 April 2022 that includes 41 pages.

Digital data processing devices:
- No. NMi-2373298-01 dated 16 December 2019 that includes 17 pages;
- No. NMi-2373298-02 dated 16 December 2019 that includes 15 pages.

Analog data processing devices:
- No. NMi-1900621-01 dated 24 May 2017 that includes 31 pages;
- No. NMi-2477319-01 dated 23 October 2020 that includes 21 pages.

Characteristics of the indicator / terminal:

Configuration	Digital load cell(s), weighing module or ADPD		
Accuracy class	Ⅱ or Ⅲ or ⅠⅢ		
Weighing ranges	Single interval Multi-interval Multiple range		
Load cell power supply	5 V DC / 12 V DC		
Fraction of the maximum permissible error	0		
Maximum number of load platforms	4		
Temperature range	-10 °C / +40 °C		
Power supply voltage	100 – 240 V AC 50/60 Hz		
Software identification	**Software module**	**Version[(1)]**	**Unique/Thumbprint ID**
	Boot service	V1.1.3 or v1.1.3 V2.0.0 or 2.0.0 V3.0.xx or 3.0.xx V5.yy.xxx or 5.yy.xxx	B645 EE8D B827409295 B827409295
	Scale server	V1.y.xx or v1.y.xx V2.y.xx or 2.y.xx 3.y.xx 5.yy.xxx	-
	ScaleLock	V1.1.xx	-
	Scale module 'IDNet or Rainbow'	V1.1.xx	-

Software identification	Scale module 'POWERCELL'	V1.x.yy V4.x.yy	-

Remark: 1, The letters x, y and z in version number are not relevant for legal metrology and can be represented by numbers or letters.

Software:
- The identification number will be displayed after pressing the following button:

    - 'Classic' version:

    - 'Pro' version and 'Pro' WPF version:

Revision History

This revision replaces the previous versions.

Revision	Date	Changes
0	2020-06-29	Initial issue.
1	2020-12-24	Changed boot service ID.
2	2024-02-29	Adding a new version of Boot service and Scale server. Correction of POWERCELL version format.
3	2025-02-18	Adding new software version for POWERCELL board. Adding report NMi-2657900-01.