



OIML Certificate





Number R76/2006-A-NL1-20.27 revision 1 Project number 3513257 Page 1 of 3

Issuing authority NMi Certin B.V.

Person responsible: M.Ph.D. Schmidt

Applicant and Mettler Toledo GmbH Manufacturer Im Langacher 44

8606 Greifensee **Switzerland**

Identification of the

certified type

An Indicator or Terminal

Type

IND570

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class II or III or III

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.



Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1 31 October 2022



Certification Board

at www.oiml.org

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.







NMi Certin B.V. The Netherlands T +31 88 6362332 certin@nmi.nl www.nmi.nl









OIML Member State

The Netherlands



Number R76/2006-A-NL1-20.27 revision 1 Project number 3513257 Page 2 of 3

OIML Certificate

The conformity was established by the results of tests and examinations provided in the associated **OIML Reports:**

- No. NMi-13200606-01 dated 17 April 2014 that includes 53 pages;
- No. NMi-13200606-02 dated 17 April 2014 that includes 17 pages;
- No. NMi-13200606-03 dated 9 October 2015 that includes 9 pages;
- No. NMi-13200606-04 dated 9 October 2015 that includes 13 pages; No. NMi-13200606-05 dated 9 October 2015 that includes 10 pages;
- No. NMi-15200584-01 dated 16 June 2016 that includes 24 pages;
- No. NMi-2372754-01 dated 12 December 2019 that includes 9 pages;
- No. NMi-2395635-01 dated 29 June 2020 that includes 20 pages.

Characteristics of the indicator:

Configuration		Analog load cells	Analog load cells with ISB	Digital load cells or weighing modules
Accuracy class	OIML R 76	Or Or		II),III) or IIII)
	OIML R 51	Y(a) or Y(b) XIII(x) or XIIII(x)		Y(II), Y(a) or Y(b) XII(x), XIII(x) or XIIII(x)
	OIML R 61		Ref (0,2)	
Maximum number of verification scale intervals		10000		-
Load cell excitation voltage		10 V DC	2,5 V DC	-
Load cell power supply voltage		-		12 V or 24 V DC
Minimum input voltage per verification scale interval		0,3 μV	0,22 μV	-
Minimum load cell resistance		29 Ω	87 Ω	-
Maximum load cell resistance		1236 Ω	3150 Ω	-
Fraction of the maximum permissible error		0,5		0
Load cell connection		6-wire		-
Maximum value of the cable length per cross wire section between the instrument and the junction box or load cells		482 m/mm²	395 m/mm²	-
Weighing range(s)			Single interval Multi-interval Multiple range	
	temperature range	-10 °C / +40 °C	+5 °C / +40 °C	-10 °C / +40 °C
Climatic — environment —	humidity	Non-condensing		
_	intended location			







OIML Certificate





Number R76/2006-A-NL1-20.27 revision 1 Project number 3513257 Page 3 of 3

Electromagnetic environment class	E2
Power supply voltage	100 – 240 V AC 50/60 Hz 24 V DC
Application	Intended to be used for direct sales to the public

Software identification:

Description	Version	Remarks		
Standard	1.xx.yyyy 2.xx.yyyy 3.xx.yyyy	For analog, IDNet and SICSpro scale types. Version 3.xx.yyyy adds hysteresis compensation function only for analog scale type.		
Standard	1.xx.yyyy 2.xx.yyyy 3.xx.yyyy	For POWERCELL scale types. Version 2.xx.yyyy adds recall of load cell firmware version; Version 3.xx.yyyy adds Hysteresis Compensation function.		
Fill-570	1.xx.yyyy	Only for automatic gravimetric filling instruments, but not mandatory		
Motion-570	2.xx.yyyy	Only for catchweighing instruments weighing dynamically during automatic operation.		

(xx is a number between 00 and 99 representing major updates of the non legally relevant part of the software and yyyy is a number between 0000 and 9999 and represents minor updates of the non legally relevant part of the software)

Revision History



This revision replaces the previous version.

Revision	Date	Change(s)
0	2020-06-29	Initial issue.
1	2022-10-31	Adding hysteresis compensation function and new Profibus option board.