M	C	OIML Certificate
OIML Member State The Netherlands		Number R76/2006-A-NL1-21.01 revision 1 Project number 3633002 Page 1 of 3
Issuing authority	NMi Certin B.V. Person responsible: M.Ph.D. Sc	hmidt
Applicant and Manufacturer	Shanghai Teraoka Electronic C No. 6058 of Nan Ting Road Ting Lin Town, Jin Shan Distric Shanghai China	
Identification of the certified type	An Indicator Type	: DIX-3000 DIX-3000SS
Characteristics	See next page	

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class (III) or (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Reports is not permitted, although either may be reproduced in full.

Issuing Authority



NMi Certin B.V. Thijsseweg 11 2629 JA Delft The Netherlands T +31 88 6362332 certin@nmi.nl www.nmi.nl NMi Certin B.V., OIML Issuing Authority NL1 14 March 2023

Certification Board

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.







OIML Member State

The Netherlands

_

OIML Certificate



Number R76/2006-A-NL1-21.01 revision 1 Project number 3633002 Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Reports:

- No. NMi-2434284-01 dated 22 October 2020 that includes 55 pages;
- No. NMi-2551360-01 dated 11 January 2021 that includes 15 pages;
- No. NMi-3633002-01 dated 14 March 2023 that includes 18 pages.

Characteristics of the indicator:

Configuration				Analog load cells	
Accuracy class					
Weighing range(s)				Single interval Multi-interval	
Maximum number of scale intervals (one weighing range)			n ≤ 7500		
Maximum number of scale intervals (multi-interval)			n ≤ 7500 (per partial weighing range)		
Maximum number partial weighing r			2		
Tare			$T \leq -Max_1 + e_1$ for multi-interval instruments		
Load cell excitation voltage			5 V DC		
Minimum signal ir	nput voltage		U _{min} = 0 mV		
Minimum input vo scale interval	oltage per verificatior	ı	0,66 µV		
Minimum load cell resistance			85 Ω		
Maximum load ce	ll resistance		3300 Ω		
Fraction of the ma	aximum permissible e	rror	0,5		
Load cell interface			6-wire (remote sensing)		
	f the cable length per between the indicato or load cells		Length 406 m/mm ² In case 4-wire connection is used the load cells are connected directly without junction box		
Temperature rang	е		-10 °C / +40 °C		
Power supply volt	age		12 V DC supplied by 100 – 240 V AC 50/60 Hz AC/DC adapter, 7,4 V internal battery		
Software identification	Filename	V	ersion	Checksum	
STE77 software	libwm.so		1.1	0142994b15a6d3a0832b87f19ca29d9b	
(1				





Number R76/2006-A-NL1-21.01 revision 1 Project number 3633002 Page 3 of 3

Revision History

This revision replaces the previous version.

)	Revision	Date	Change(s)		
	0	2021-01-12	Initial issue		
	1	2023-03-14	Addition of alternative AC/DC switch power adapter		

÷