

OIML Member State
The Netherlands

Number R76/2006-A-NL1-21.24 revision 3
Project number 3493907
Page 1 of 3

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and
Manufacturer

Bizerba SE & Co. KG
Wilhelm-Kraut-Str. 65
72336 Balingen
Germany

Identification of the
certified type

A Non-automatic weighing instrument
Type

: Q1...100... (bench type)
Q1...100... (pole type)
Q1...400... (hanging type)
Q1...800... (pole type)
'...' denotes a placeholder for
alternative variants not affecting
metrological parameters.

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
22 September 2023

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



OIML Member State
The Netherlands

Number R76/2006-A-NL1-21.24 revision 3
Project number 3493907
Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated report(s):

- No. NMI-2477454-01 revision 1 dated 15 September 2023 that includes 51 pages;
- No. NMI-2477454-02 dated 15 April 2021 that includes 18 pages;
- No. NMI-2529842-01 dated 15 April 2021 that includes 17 pages;
- No. NMI-3001387-01 dated 18 October 2022 that includes 26 pages;
- No. NMI-3001387-02 dated 18 October 2022 that includes 14 pages.

Characteristics of the non-automatic weighing instrument:

| | | |
|--|--|---|
| Accuracy class | (III) | |
| Maximum capacity | Weighing module QBase | $3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$ |
| | Weighing module ScalePad USB | $\text{Max} \leq 15 \text{ kg}$ |
| | Weighing module ADC-Box USB | Calculated using the compatibility of modules form, contained in EN 45501:2015 clause F.4 |
| Verification scale interval | $e \geq 1 \text{ g}$ | |
| Weighing ranges | Single interval Multi-interval | |
| Maximum number of scale intervals | $n \leq 6000$ | |
| Maximum number of scale intervals (multi-interval) | $n \leq 3000$ (per partial weighing range) | |
| Maximum number of partial weighing ranges | 2 | |
| Semi-automatic Tare | $T \leq -100\% \text{ of Max}$ | |
| Preset Tare | 100% of Max; 100% of Max ₁ , in case of multi-interval instruments | |
| Temperature range | -10 °C / +40 °C | |
| Power supply voltage | 120 – 240 V AC 50/60 Hz | |
| Application | Intended to be used for direct sales to the public | |

Terminal software identification:

| Scale OEM Module | Software version number (legally relevant) ¹⁾ | Software ID (Windows) | Software ID (Linux) |
|------------------|--|-----------------------|---------------------|
| | 013 | 7640 | 0369 |
| | 015 | 8282 | 1171 |

OIML Member State
The Netherlands

Number R76/2006-A-NL1-21.24 revision 3
Project number 3493907
Page 3 of 3

A/D board software identification:

| | Version number (legally relevant) ¹⁾ | Software ID |
|-----|---|-------------|
| ADW | 010 | 6153 |
| | 011 | 5232 |
| | 030 | 1446 |

¹⁾ The identification of the non-legally relevant software version can optionally be added to the identification of the legally relevant software version, separated by "::" (example: "010::x.y" or "015::x.y" or "030::x.y").

Revision History

This revision replaces the previous versions.

| Revision | Date | Change(s) |
|----------|------------|---|
| 0 | 2021-04-15 | Initial issue |
| 1 | 2022-10-18 | Adding Q1...400... (hanging type) and Q1...100.... (pole type). Adding ScalePad USB and ADC-Box USB. |
| 2 | 2023-01-05 | Addition of a new version. |
| 3 | 2023-09-22 | Replace type evaluation report NMI-2477454-01 with NMI-2477454-01 revision 1. |