

OIML Member State
The Netherlands

Number R76/2006-A-NL1-22.07 revision 1
Project number 3513725
Page 1 of 2

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and
Manufacturer

Teraoka Seiko Co.,Ltd.
5-13-12 Kugahara, Ohta-Ku
Tokyo 146-8580
Japan

Identification of the
certified type

A Terminal
Type : SM-6000 Printer Console

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
15 July 2022

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated OIML Report:

- No. NMI-2733270-03 dated 25 February 2022 that includes 22 pages.

Characteristics of the terminal:

Configuration	connection to a non-automatic weighing instrument intended to be used for direct sales to the public		
Maximum number of connected non-automatic weighing instruments	128 units		
Temperature range	-10 °C / +40 °C		
Power supply voltage	24 V DC supplied by an AC/DC adapter		
Application	Intended to be used for direct sales to the public		
Software identification		hardware sealing	software sealing
	Software version number	24.xx.xx-xx (xx is a number between 00 and 99)	
	HS SBS DRIVER version: checksum:	not applicable not applicable	11.1.1 BA5A6E5C

Revision History

This revision replaces the previous version.

Revision	Date	Change(s)
0	28 February 2022	Initial issue.
1	15 July 2022	New software identification