

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-22.14 revision 0  
Project number 2587354  
Page 1 of 2

Issuing authority

NMi Certin B.V.  
Person responsible: M.Ph.D. Schmidt

Applicant and  
Manufacturer

Fieldpiece Instruments Inc.  
1636W Collins Ave,  
Orange, CA 92867  
United States of America

Identification of the  
certified type

A **Non-automatic weighing instrument**  
Type : SRS3EC

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76-1:2006** for accuracy class III

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

**NMi Certin B.V., OIML Issuing Authority NL1**  
8 July 2022

Certification Board

NMi Certin B.V.  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Report(s):

- No. NMI-2587354-01 dated 8 July 2022 that includes 51 pages.

## Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	Max $\leq$ 100 kg
Verification scale interval	e $\geq$ 50 g
Weighing range	Single interval
Maximum number of scale intervals	n $\leq$ 2000
Tare	T $\leq$ -Max
Temperature range	0 °C / 50 °C
Power supply voltage	Platform: 6,3-9V DC through batteries Remote display: 9V DC through batteries
Software identification	Version number
	1.017

The software identification is displayed after turning on the remote and holding the power button for 5 seconds.

## Revision History

Revision	Date	Change(s)
0	2022-07-08	Initial issue.