

OIML Member State
The Netherlands

Number R76/2006-A-NL1-22.17 revision 0
Project number 2663177
Page 1 of 2

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and
Manufacturer

Shenzhen Comen Medical Instruments Co., Ltd.
Floor 10, Floor 11 and Section C of Floor 12 of building 1A
& Floor 1 to Floor 5 of Building 2, FIYTA Timepiece Building,
Nanhuan Avenue, Matian Sub-district
Guangming District, Shenzhen
518106 Guangdong
P.R. China

Identification of the
certified type

A Non-automatic weighing instrument
Type : Neonatal Incubator B8, B8A, B6, B6A, B5, B3, B3A,
B2, B2A, B2B
Infant Radiant Warmer BQ80, BQ80A

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
26 August 2022

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Reports:

- No. NMI-2663177-01 dated 26 August 2022 that includes 46 pages;
- No. NMI-2663177-02 dated 26 August 2022 that includes 20 pages;
- No. NMI-2663177-03 dated 26 August 2022 that includes 14 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	(III)	
Maximum capacity	Max = 8 kg	
Verification scale interval	e = 10 g	
Weighing range	Single interval	
Maximum number of scale intervals (one weighing range)	n ≤ 800 divisions	
Tare	T ≤ 3 kg	
Temperature range	+10 °C / +40 °C	
Power supply voltage	200 – 240 V AC 50/60 Hz Battery 12 V DC	
Application	To be used only for weighing patients	
Software identification	Version number	Neonatal Incubator V1.1.7.x (x = 0... 99)
		Infant Radiant Warmer V1.0.7.x (x = 0... 99)

Revision History

Revision	Date	Change(s)
0	2022-08-26	Initial issue.