

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-23.02 revision 0  
Project number 3609407  
Page 1 of 2

Issuing authority

NMi Certin B.V.  
Person responsible: M.Ph.D. Schmidt

Applicant and  
Manufacturer

SHANGHAI ISHIDA ELECTRONIC SCALES CO., LTD.  
Building 2, No.86, Min Xue Road,  
Pudong, Shanghai, 201209  
China

Identification of the  
certified type

**A Non-automatic weighing instrument**  
Type : S-boxM series

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76-1:2006** for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

**NMi Certin B.V., OIML Issuing Authority NL1**  
27 January 2023

Certification Board

NMi Certin B.V.  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Reports:

- No. NMI-2399622-01 dated 11 March 2020 that includes 45 pages;
- No. NMI-2399622-02 dated 11 March 2020 that includes 28 pages;
- No. NMI-2486567-01 dated 7 January 2021 that includes 45 pages;
- No. NMI-2486567-02 dated 7 January 2021 that includes 18 pages.

### Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
Verification scale interval	$e \geq 1 \text{ g}$
Weighing range	Single interval
Maximum number of scale intervals	$n \leq 3000$
Temperature range	$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Power supply voltage	9 V DC through plug-in power supply; 6 V DC with internal rechargeable battery; 7,5 V DC with dry battery
Software identification	2.yy.xxxx (y, x = 0... 9)

The software identification is displayed at start-up.

### Revision History

Revision	Date	Change(s)
0	27 January 2023	Initial issue