

OIML Member State
The Netherlands

Number R76/2006-A-NL1-23.04 revision 0
Project number 2752431
Page 1 of 2

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and
Manufacturer

Zebra Technologies Corp
1 Zebra Plaza
Holtsville, NY 11742
United States of America

Identification of the
certified type

A Non-automatic weighing instrument
Type : MP70X1 single interval
MP70X2 multi-interval

'X' denotes alternative variants not
affecting metrological parameters.

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

This certificate and supporting reports comply with the requirements of OIML-CS-PD-07 clause 6.2.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
17 February 2023

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated OIML Reports:

- No. SN 1379 dated 12 May 2017 that includes 47 pages;
- No. P02070 dated 12 May 2017 that includes 14 pages;
- No. P02848 dated 8 October 2020 that includes 16 pages;
- No. NMI-2752431-01 dated 17 February 2023 that includes 45 pages;
- No. NMI-2752431-02 dated 17 February 2023 that includes 20 pages.

Characteristics of the non-automatic weighing instrument:

| | |
|--|--|
| Accuracy class | III |
| Maximum capacity | Max \leq 15 kg |
| Verification scale interval | $e \geq$ 2 g |
| Weighing ranges | Single interval Multi-interval |
| Maximum number of scale intervals (one weighing range) | $n \leq$ 3000 |
| Maximum number of scale intervals (multi-interval) | $n \leq$ 3000 (per partial weighing range) |
| Maximum number of partial weighing ranges | 2 |
| Temperature range | 0 °C / +40 °C |
| Power supply voltage | 100 – 240 V AC 50/60 Hz 12 V DC |
| Application | Intended to be used for direct sales to the public |
| Software identification | Version number 1.04F |

When pressing and holding the zero button (<0>) for a period of at least 3 seconds at stable condition, the operator display will display:

- Software identification;
- Calibration counter (in the form "C xxx");
- Parameter counter (in the form "P xxx").

Revision History

| Revision | Date | Change(s) |
|----------|------------|----------------|
| 0 | 2023-02-17 | Initial issue. |