

OIML Member State
The Netherlands

Number R76/2006-A-NL1-23.12 revision 0
Project number 3653225
Page 1 of 3

Issuing authority NMI Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and Manufacturer Grupo Epelsa, S.L.U.
C/ Punto Net, 3
28805 Alcalá de Henares (Madrid)
Spain

Identification of the certified type **A Non-automatic weighing instrument**
Type : BP-XX

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class **III** or **IIII**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

This certificate and supporting reports comply with the requirements of OIML-CS-PD-07 clause 6.2.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Reports) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**
4 May 2023

Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated reports:

Module	Type	Report	Issue date	Pages
Load cell	ZSF-A-...	R60/2000-NL1-05.17A	12 October 2005	40
	ZSFY-A-...	R60/2000-NL1-05.17B	11 October 2005	37
Indicator	Energy	NMi-3478944-01	11 March 2022	9
Indicator	ORION PLUS CYBER PLUS	NMi-13200261-01	16 September 2013	49
Indicator	ML-500 CT-12	NMi-2535139-01 revision 1	31 January 2023	29
		NMi-2535139-02 revision 1	31 January 2023	21

For each model the compatibility of load cells and indicator is established by the manufacturer by means of the compatibility of modules form, contained in OIML R76 edition 2006 Annex F.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III or IIII
Maximum capacity	30 t ≤ Max ≤ 60 t
Verification scale interval	e ≥ 10 kg
Weighing ranges	Single interval Multi-interval Multiple range
Maximum number of scale intervals (Single interval)	n ≤ 3000 divisions
Maximum number of scale intervals	n ≤ 3000 divisions (per weighing range or partial weighing range)
Maximum number of partial weighing ranges	2
Load cell excitation voltage	5 V DC
Minimum input voltage per verification scale interval	0,66 μV
Minimum load cell resistance	35 Ω
Maximum load cell resistance	1100 Ω ± 50 Ω
Fraction of the maximum permissible error	0,5
Load cell interface	6-wire with sense technology
Maximum number of load platforms	2
Temperature range	-10 °C / +40 °C
Power supply voltage	220 – 230 V AC 50/60 Hz

Software identification	Energy		Version number: En.XX.X.XX 1X (X = 0...9 or X = A...Z)
	ORION PLUS CYBER PLUS		XXXX.Y.XXX. 10 or XXXX.X.XXX. 0X (where X is 0...9 or A...Z) (where Y is 0...9 or A...Z) The '0' or '10' indicates the metrological version of the software.
	ML-500 CT-12	Firmware	MBPT-CCII-000Y00-0.xx, where Y → 1= Mini BPTC board & CS5530 board 2=A+M board: CC = country, II = language (variable), 0.xx = version, xx= variable.
		ScalePos (Windows)	Version: 1.0.2.3 Checksum: E3E57643A0178BE8C21F9CFF53BFB80D
		TouchScale (Linux)	Version: 1 Checksum: dcf11d09b220209e7f92ef48494d6dc2
		TouchScale (windows)	Version: 1 Checksum: 9669792b564f793471c962c02bd8d912

Revision History

Revision	Date	Change(s)
0	2023-05-04	Initial issue.