

OIML Member State
The Netherlands

Number R76/2006-A-NL1-23.14 revision 0
Project number 3598849
Page 1 of 2

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and
Manufacturer

Técnicas de Electrónica y Automatismos, S.A.
C/Espronceda 180
08018 Barcelona
Spain

Identification of the
certified type

An **Indicator**
Type : SMART DIGITAL series

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class (III) or (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
30 May 2023

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



OIML Member State
The Netherlands

Number R76/2006-A-NL1-23.14 revision 0
Project number 3598849
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-3598849-01 dated 30 May 2023 that includes 22 pages;
- No. NMI-3598849-02 dated 30 May 2023 that includes 12 pages;
- No. NMI-3598849-03 dated 30 May 2023 that includes 18 pages.

Characteristics of the indicator:

Configuration	Digital load cells
Accuracy class	Ⓜ or ⓂⓂ
Weighing range(s)	Single interval Multiple range
Maximum number of weighing ranges	2
Load cell power supply voltage	15 V DC
Fraction of the maximum permissible error	0
Power supply voltage	100 – 240 V AC 50/60 Hz
Software identification	Version number: 1.xxx (xxx is a number between 000 and 999)

Software:

- The identification number will be displayed at start-up;
- The indicator has embedded software.

Revision History

Revision	Date	Change(s)
0	30-05-2023	Initial issue.