

OIML Member State
The Netherlands

Number R76/2006-A-NL1-23.24 revision 0
Project number 3726147
Page 1 of 2

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Manufacturer

Teraoka Seiko Co. Ltd.
13-12 Kugahara, 5-Chome
Ohta-Ku, Tokyo 146-8580
Japan

Identification of the certified type

A Non-automatic weighing instrument
Type : DS-502HD (BT)

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
21 December 2023

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



OIML Member State
The Netherlands

Number R76/2006-A-NL1-23.24 revision 0
Project number 3726147
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-3596436-01 revision 1 dated 16 March 2023 that includes 47 pages;
- No. NMI-3726147-01 dated 21 December 2023 that includes 33 pages;
- No. NMI-3726147-02 dated 21 December 2023 that includes 8 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$30 \text{ kg} \leq \text{Max} \leq 300 \text{ kg}$
Verification scale interval	$e \geq 10 \text{ g}$
Weighing range(s)	Single interval Multi-interval
Maximum number of scale intervals (one weighing range)	$n \leq 3000$
Maximum number of scale intervals (multi-interval)	$n \leq 3000$ (per partial weighing range)
Maximum number of partial weighing ranges	2
Tare	$T \leq -50\%$ for instruments with one weighing range $T \leq -\text{Max}_1$ for multi-interval instruments
Temperature range	$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Power supply voltage	9 V DC battery
Software identification Version number	v 1,xx (xx=00.....99)

Revision History

Revision	Date	Change(s)
0	2023-12-21	Initial issue.