

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-24.09 revision 1  
Project number 3831307  
Page 1 of 2

Issuing authority

NMi Certin B.V.  
Person responsible: M.Ph.D. Schmidt

Applicant and  
Manufacturer

Anytime Postage LLC  
2803 Philadelphia Pike  
Suite B #1575  
Claymont, DE 19703  
United States of America

Identification of the  
certified type

**A Non-automatic weighing instrument**  
Type : PS-15

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76-1:2006** for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

**NMi Certin B.V., OIML Issuing Authority NL1**  
5 September 2024

Certification Board

NMi Certin B.V.  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



**OIML Member State**  
The Netherlands


Number R76/2006-A-NL1-24.09 revision 1  
Project number 3831307  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-3654085-01 dated 16 April 2024 that includes 25 pages;
- No. NMI-3654085-02 dated 16 April 2024 that includes 20 pages;
- No. NMI-3654085-03 dated 16 April 2024 that includes 20 pages.

### Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	Max = 15 kg
Minimum capacity	Min = 100 g
Verification scale interval	e = 5 g
Weighing range	Single interval
Maximum number of scale intervals	n = 3000
Temperature range	0 °C / + 40 °C
Power supply voltage	100 V – 240 V AC 50/60 Hz
Application	Intended to be used for determining a conveying tariff Intended to be used for direct sales to the public
Software identification weighing module (Firmware)	V3.012
Software identification Scale SDK	V1.014
Software Identification Terminal (App)	Vz.y.xxx (non-legally relevant)
Software identification Terminal (Metrological)	V2.030

The software identification is displayed in the information screen after pressing .

### Revision History

This revision replaces the previous version.

Revision	Date	Change(s)
0	2024-04-16	Initial issue.
1	2024-09-05	Correction of the software identifications.